



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109004

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/04/2020	14:05
2	L2464-05		1	8	100.00	>212.0	>212.0	05/04/2020	14:32
3	L2464-05DUP		1	8	100.00	>212.0	>212.0	05/04/2020	14:50
4	L2464-07		1	9	100.00	>212.0	>212.0	05/04/2020	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 109004

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp5420 WorkList ID : 138440 Department : Wet-Chemistry Date : 05-04-2020 13:15:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/08/2020	Water	L2464-05	Flash Point	Cool 4 deg C	PSEG01	L23	TRE-50028	05/01/2020	1010A
05/08/2020	Water	L2464-07	Flash Point	Cool 4 deg C	PSEG01	L23	TRE-50029	05/01/2020	1010A

Date/Time 05/04/20 13:25
Received by: KS
Relinquished by: JP

Date/Time 05/04/20 15:25
Received by: JP
Relinquished by: KS