


LB 109006

INSTRUMENT ID: CY2

7471B

 μ Abs.:

57010



Linear

A= 0.0000e+000

B= 1.7392e-004

C= 1.1188e-002

Rho= 0.9998852

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	33	0.000	33				
0.2	0.200	0.162	-0.038	868	0.0 %	868				
2.5	2.500	2.457	-0.043	14062	0.0 %	14062				
5.0	5.000	5.080	0.080	29144	0.0 %	29144				
7.5	7.500	7.558	0.058	43393	0.0 %	43393				
10.0	10.000	9.926	-0.074	57010	0.0 %	57010				

LB109006 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 May 2020 09:13:34

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	33	-	0.00007471B	PPB		05 May 2020 09:42:12	S	Std
0.2 - 1	50.2	868	-	0.20007471B	PPB		05 May 2020 09:44:28	S	Std
2.5 - 1	52.5	14062	-	2.50007471B	PPB		05 May 2020 09:46:45	S	Std
5.0 - 1	55	29144	-	5.00007471B	PPB		05 May 2020 09:49:02	S	Std
7.5 - 1	57.5	43393	-	7.50007471B	PPB		05 May 2020 09:51:18	S	Std
10.0 - 1	59	57010	-	10.00007471B	PPB		05 May 2020 09:53:35	S	Std
ICV31 - 1	ICV31	20783	3.6257	-7471B	PPB		05 May 2020 09:56:41	U	SMPL
ICB31 - 1	ICB31	-97	-0.0057	-7471B	PPB		05 May 2020 09:58:55	U	SMPL
CCV48 - 1	CCV48	26917	4.6925	-7471B	PPB		05 May 2020 10:01:11	U	SMPL
CCB48 - 1	CCB48	-103	-0.0067	-7471B	PPB		05 May 2020 10:03:25	U	SMPL
CRA - 1	CRA	829	0.1554	-7471B	PPB		05 May 2020 10:05:41	U	SMPL
HighStd - 1	HighStd	55657	9.6909	-7471B	PPB		05 May 2020 10:07:56	U	SMPL
ChkStd - 1	ChkStd	38439	6.6964	-7471B	PPB		05 May 2020 10:10:11	U	SMPL
PB128654BL - 1	PBS	-107	-0.0074	-7471B	PPB		05 May 2020 10:12:29	U	SMPL
PB128654BS - 1	LCSS	20918	3.6492	-7471B	PPB		05 May 2020 10:14:45	U	SMPL
L2481-01 - 1	HSM-01-W	182234	31.7049	-7471B	PPB		05 May 2020 10:17:02	U	SMPL
L2481-02 - 1	HSM-01-R	622109	HIGH	-7471B	PPB		05 May 2020 10:19:18	X	
L2481-03 - 1	HSM-01-C	461668	HIGH	-7471B	PPB		05 May 2020 10:21:14	X	
L2481-04 - 1	HSM-02-W	295344	51.3768	-7471B	PPB		05 May 2020 10:23:27	U	SMPL
L2481-05 - 1	HSM-02-R	1063334	HIGH	-7471B	PPB		05 May 2020 10:25:54	X	
CCV49 - 1	CCV49	31345	5.4626	-7471B	PPB		05 May 2020 10:34:32	U	SMPL
CCB49 - 1	CCB49	-235	-0.0297	-7471B	PPB		05 May 2020 10:36:48	U	SMPL
L2481-06 - 1	HSM-02-C	476364	HIGH	-7471B	PPB		05 May 2020 10:39:05	X	
L2481-07 - 1	HSM-03-W	174355	30.3346	-7471B	PPB		05 May 2020 10:40:54	U	SMPL
L2481-08 - 1	HSM-03-R	555425	HIGH	-7471B	PPB		05 May 2020 10:43:18	X	
L2481-09 - 1	HSM-03-C	378331	HIGH	-7471B	PPB		05 May 2020 10:45:13	X	
L2481-09DUP - 1	HSM-03-CDUP	370167	HIGH	-7471B	PPB		05 May 2020 10:47:30	X	
L2481-09MS - 1	HSM-03-CMS	401754	HIGH	-7471B	PPB		05 May 2020 10:49:31	X	
L2481-09MSD - 1	HSM-03-CMSD	369492	HIGH	-7471B	PPB		05 May 2020 10:51:29	X	
L2481-10 - 1	P-03-R	38553	6.7162	-7471B	PPB		05 May 2020 10:53:24	U	SMPL
L2481-11 - 1	HSM-04-W	438691	HIGH	-7471B	PPB		05 May 2020 10:55:50	X	
L2481-12 - 1	HSM-04-R	786987	HIGH	-7471B	PPB		05 May 2020 10:57:42	X	
CCV50 - 1	CCV50	29156	5.0819	-7471B	PPB		05 May 2020 10:59:44	U	SMPL
CCB50 - 1	CCB50	-421	-0.0620	-7471B	PPB		05 May 2020 11:02:39	U	SMPL
L2481-13 - 1	HSM-04-C	332753	57.8829	-7471B	PPB		05 May 2020 11:04:56	U	SMPL
L2481-14 - 1	HSM-05-W	136755	23.7953	-7471B	PPB		05 May 2020 11:07:11	U	SMPL
L2481-15 - 1	HSM-05-R	722463	HIGH	-7471B	PPB		05 May 2020 11:09:44	X	
L2481-16 - 1	HSM-05-C	574073	HIGH	-7471B	PPB		05 May 2020 11:11:35	X	
L2481-17 - 1	HSM-06-W	355816	61.8939	-7471B	PPB		05 May 2020 11:13:58	U	SMPL
L2481-18 - 1	HSM-06-R	863193	HIGH	-7471B	PPB		05 May 2020 11:16:28	X	
L2481-19 - 1	HSM-06-C	269553	46.8913	-7471B	PPB		05 May 2020 11:18:36	U	SMPL
L2481-20 - 1	P-03-C	9890	1.7312	-7471B	PPB		05 May 2020 11:21:43	U	SMPL
PB128655BL - 1	PBS	-221	-0.0272	-7471B	PPB		05 May 2020 11:24:16	U	SMPL
PB128655BS - 1	LCSS	20669	3.6059	-7471B	PPB		05 May 2020 11:26:34	U	SMPL
CCV51 - 1	CCV51	30685	5.3479	-7471B	PPB		05 May 2020 11:28:49	U	SMPL
CCB51 - 1	CCB51	-196	-0.0229	-7471B	PPB		05 May 2020 11:31:05	U	SMPL
L2482-01 - 1	HSA-01-R	1141811	HIGH	-7471B	PPB		05 May 2020 11:33:22	X	
L2482-03 - 1	HSA-01-C	317849	HIGH	-7471B	PPB		05 May 2020 11:35:10	X	
L2482-03DUP - 1	HSA-01-CDUP	309509	HIGH	-7471B	PPB		05 May 2020 11:38:43	X	
L2482-03MS - 1	HSA-01-CMS	323747	HIGH	-7471B	PPB		05 May 2020 11:44:00	X	
L2482-03MSD - 1	HSA-01-CMSD	322180	HIGH	-7471B	PPB		05 May 2020 11:45:54	X	
L2482-05 - 1	HSA-02-R	1085475	HIGH	-7471B	PPB		05 May 2020 11:47:50	X	
L2482-07 - 1	HSA-02-C	474157	HIGH	-7471B	PPB		05 May 2020 11:49:45	X	
L2482-09 - 1	LP-01-R	358382	HIGH	-7471B	PPB		05 May 2020 11:53:09	X	
L2482-11 - 1	LP-01-C	55743	9.7059	-7471B	PPB		05 May 2020 11:55:13	U	SMPL
L2482-13 - 1	LP-02-R	694773	HIGH	-7471B	PPB		05 May 2020 11:57:41	X	
CCV52 - 1	CCV52	29155	5.0818	-7471B	PPB		05 May 2020 11:59:32	U	SMPL
CCB52 - 1	CCB52	-639	-0.0999	-7471B	PPB		05 May 2020 12:02:15	U	SMPL
L2482-15 - 1	LP-02-C	20305	3.5426	-7471B	PPB		05 May 2020 12:04:32	U	SMPL
L2482-17 - 1	LP-03-R	360543	HIGH	-7471B	PPB		05 May 2020 12:06:48	X	
L2482-19 - 1	LP-03-C	108266	18.8406	-7471B	PPB		05 May 2020 12:08:40	U	SMPL
L2482-21 - 1	LW-01-R	3487	0.6176	-7471B	PPB		05 May 2020 12:11:00	U	SMPL
L2482-23 - 1	LW-01-C	5308	0.9343	-7471B	PPB		05 May 2020 12:13:19	U	SMPL
L2482-25 - 1	LW-02-R	2498	0.4456	-7471B	PPB		05 May 2020 12:15:35	U	SMPL
L2482-27 - 1	LW-02-C	1435	0.2608	-7471B	PPB		05 May 2020 12:17:52	U	SMPL
L2482-29 - 1	LW-03-R	1607	0.2907	-7471B	PPB		05 May 2020 12:20:09	U	SMPL
L2482-31 - 1	LW-03-C	3749	0.6632	-7471B	PPB		05 May 2020 12:22:26	U	SMPL
L2482-33 - 1	P-01-R	1506	0.2731	-7471B	PPB		05 May 2020 12:24:42	U	SMPL
CCV53 - 1	CCV53	30511	5.3176	-7471B	PPB		05 May 2020 12:26:58	U	SMPL
CCB53 - 1	CCB53	-168	-0.0180	-7471B	PPB		05 May 2020 12:29:13	U	SMPL
L2482-35 - 1	P-01-C	947	0.1759	-7471B	PPB		05 May 2020 12:31:30	U	SMPL
L2482-37 - 1	P-02-R	1170	0.2147	-7471B	PPB		05 May 2020 12:33:46	U	SMPL
L2482-39 - 1	P-02-C	387	0.0785	-7471B	PPB		05 May 2020 12:36:02	U	SMPL
L2481-09LX5 - 1		116202	20.2208	-7471B	PPB		05 May 2020 12:38:18	U	SMPL
L2481-09A - 1		379205	HIGH	-7471B	PPB		05 May 2020 12:40:34	X	
L2482-03LX5 - 1		93605	16.2908	-7471B	PPB		05 May 2020 12:42:28	U	SMPL
L2482-03A - 1		323010	HIGH	-7471B	PPB		05 May 2020 12:44:50	X	
CCV54 - 1	CCV54	29828	5.1988	-7471B	PPB		05 May 2020 12:46:45	U	SMPL
CCB54 - 1	CCB54	-205	-0.0245	-7471B	PPB		05 May 2020 12:49:05	U	SMPL
L2481-01DLX10 - 1		17302	3.0203	-7471B	PPB		05 May 2020 12:51:22	U	SMPL
L2481-02DLX10 - 1		124575	21.6770	-7471B	PPB		05 May 2020 12:53:38	U	SMPL

LB109006 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 May 2020 09:13:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
L2481-03DLX10 - 1		68738	11.9660		-7471B	PPB	05 May 2020 12:55:54	U	SMPL
L2481-04DLX10 - 1		28440	4.9574		-7471B	PPB	05 May 2020 12:58:13	U	SMPL
L2481-05DLX10 - 1		371587	HIGH		-7471B	PPB	05 May 2020 13:00:31	X	
L2481-06DLX10 - 1		73499	12.7940		-7471B	PPB	05 May 2020 13:02:23	U	SMPL
L2481-07DLX10 - 1		16153	2.8205		-7471B	PPB	05 May 2020 13:04:44	U	SMPL
L2481-08DLX10 - 1		108277	18.8425		-7471B	PPB	05 May 2020 13:07:03	U	SMPL
L2481-09DLX10 - 1		55974	9.7461		-7471B	PPB	05 May 2020 13:09:20	U	SMPL
L2481-09DUPDLX10 - 1		55501	9.6638		-7471B	PPB	05 May 2020 13:11:39	U	SMPL
CCV55 - 1	CCV55	30338	5.2875		-7471B	PPB	05 May 2020 13:13:58	U	SMPL
CCB55 - 1	CCB55	-138	-0.0128		-7471B	PPB	05 May 2020 13:16:17	U	SMPL
L2481-09MSDLX10 - 1		56314	9.8052		-7471B	PPB	05 May 2020 13:21:47	U	SMPL
L2481-09MSDDLX10 - 1		55349	9.6374		-7471B	PPB	05 May 2020 13:24:02	U	SMPL
L2481-09LDLX50 - 1		10651	1.8636		-7471B	PPB	05 May 2020 13:26:20	U	SMPL
L2481-09ADLX10 - 1		54453	9.4815		-7471B	PPB	05 May 2020 13:28:39	U	SMPL
L2481-11DLX10 - 1		73551	12.8030		-7471B	PPB	05 May 2020 13:30:55	U	SMPL
L2481-12DLX10 - 1		218304	37.9782		-7471B	PPB	05 May 2020 13:33:12	U	SMPL
L2481-13DLX10 - 1		35402	6.1682		-7471B	PPB	05 May 2020 13:35:31	U	SMPL
L2481-14DLX5 - 1		27544	4.8016		-7471B	PPB	05 May 2020 13:37:53	U	SMPL
L2481-15DLX10 - 1		180704	31.4388		-7471B	PPB	05 May 2020 13:40:11	U	SMPL
L2481-16DLX10 - 1		101369	17.6411		-7471B	PPB	05 May 2020 13:42:28	U	SMPL
CCV56 - 1	CCV56	29658	5.1692		-7471B	PPB	05 May 2020 13:44:51	U	SMPL
CCB56 - 1	CCB56	-211	-0.0255		-7471B	PPB	05 May 2020 13:47:10	U	SMPL
L2481-17DLX10 - 1		35688	6.2180		-7471B	PPB	05 May 2020 13:49:28	U	SMPL
L2481-18DLX10 - 1		238940	41.5671		-7471B	PPB	05 May 2020 13:51:44	U	SMPL
L2481-19DLX10 - 1		23356	4.0732		-7471B	PPB	05 May 2020 13:54:02	U	SMPL
L2482-01DLX10 - 1		570624	HIGH		-7471B	PPB	05 May 2020 13:56:32	X	
L2482-03DUPDLX10 - 1		47947	8.3500		-7471B	PPB	05 May 2020 13:58:24	U	SMPL
L2482-03MSDLX10 - 1		43371	7.5542		-7471B	PPB	05 May 2020 14:01:01	U	SMPL
L2482-03MSDDLX10 - 1		48106	8.3777		-7471B	PPB	05 May 2020 14:03:20	U	SMPL
L2482-03LDLX50 - 1		48366	8.4229		-7471B	PPB	05 May 2020 14:05:38	U	SMPL
L2482-03ADLX10 - 1		9154	1.6032		-7471B	PPB	05 May 2020 14:07:56	U	SMPL
CCV57 - 1	CCV57	47241	8.2272		-7471B	PPB	05 May 2020 14:10:14	U	SMPL
CCB57 - 1	CCB57	30445	5.3061		-7471B	PPB	05 May 2020 14:12:31	U	SMPL
L2482-05DLX10 - 1		-85	-0.0036		-7471B	PPB	05 May 2020 14:14:50	U	SMPL
L2482-07DLX10 - 1		474648	HIGH		-7471B	PPB	05 May 2020 14:17:07	X	
L2482-09DLX10 - 1		79791	13.8883		-7471B	PPB	05 May 2020 14:18:58	U	SMPL
L2482-13DLX10 - 1		58821	10.2412		-7471B	PPB	05 May 2020 14:21:22	U	SMPL
L2482-17DLX10 - 1		160036	27.8443		-7471B	PPB	05 May 2020 14:23:41	U	SMPL
L2482-19DLX5 - 1		61142	10.6449		-7471B	PPB	05 May 2020 14:26:00	U	SMPL
L2481-02DL2X50 - 1		20515	3.5791		-7471B	PPB	05 May 2020 14:28:22	U	SMPL
L2481-03DL2X25 - 1		25808	4.4997		-7471B	PPB	05 May 2020 14:30:41	U	SMPL
L2481-05DL2X100 - 1		29686	5.1741		-7471B	PPB	05 May 2020 14:32:58	U	SMPL
L2481-08DL2X50 - 1		55758	9.7085		-7471B	PPB	05 May 2020 14:38:50	U	SMPL
CCV58 - 1	CCV58	21586	3.7654		-7471B	PPB	05 May 2020 14:41:06	U	SMPL
CCB58 - 1	CCB58	30295	5.2800		-7471B	PPB	05 May 2020 14:47:23	U	SMPL
L2481-06DL2X25 - 1		-91	-0.0046		-7471B	PPB	05 May 2020 14:49:39	U	SMPL
L2481-11DL2X25 - 1		32014	5.5790		-7471B	PPB	05 May 2020 14:51:58	U	SMPL
L2481-12DL2X100 - 1		3008	0.5343		-7471B	PPB	05 May 2020 14:54:14	U	SMPL
L2481-15DL2X100 - 1		23120	4.0322		-7471B	PPB	05 May 2020 14:56:33	U	SMPL
L2481-16DL2X100 - 1		19052	3.3247		-7471B	PPB	05 May 2020 14:58:49	U	SMPL
L2481-18DL2X100 - 1		19986	3.4871		-7471B	PPB	05 May 2020 15:01:06	U	SMPL
CCV59 - 1	CCV59	28300	4.9331		-7471B	PPB	05 May 2020 15:03:23	U	SMPL
CCB59 - 1	CCB59	28439	4.9572		-7471B	PPB	05 May 2020 15:08:22	U	SMPL
L2482-01DL2X250 - 1		-146	-0.0142		-7471B	PPB	05 May 2020 15:10:38	U	SMPL
L2482-05DL2X250 - 1		46577	8.1118		-7471B	PPB	05 May 2020 15:12:57	U	SMPL
L2482-07DL2X25 - 1		32492	5.6621		-7471B	PPB	05 May 2020 15:15:13	U	SMPL
L2482-09DL2X25 - 1		31924	5.5633		-7471B	PPB	05 May 2020 15:17:31	U	SMPL
L2482-17DL2X25 - 1		22113	3.8570		-7471B	PPB	05 May 2020 15:19:50	U	SMPL
L2482-13DL2X50 - 1		25762	4.4917		-7471B	PPB	05 May 2020 15:36:36	U	SMPL
CCV60 - 1	CCV60	35152	6.1247		-7471B	PPB	05 May 2020 15:39:34	U	SMPL
CCB60 - 1	CCB60	30079	5.2425		-7471B	PPB	05 May 2020 15:42:10	U	SMPL
		-167	-0.0179		-7471B	PPB	05 May 2020 15:44:27	U	SMPL