



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 5/6/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 16:24  
In Date: 05/05/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 05/06/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLIDS-OVEN

QC:LB109013

| Lab ID   | Client SampleID       | Dish # | Dish Wt (g) (A) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------------|--------|-----------------|--------------------------|----------------------------|---------|----------|
| L2401-09 | DRAINAGE-SEWALE       | 1      | 1.13            | 9.87                     | 8.94                       | 89.4    |          |
| L2401-10 | DRAINAGE-SEWALE       | 2      | 1.18            | 9.51                     | 8.83                       | 91.8    |          |
| L2483-02 | 6FT-TP-1              | 3      | 1.16            | 9.62                     | 8.6                        | 87.9    |          |
| L2483-04 | 6FT-TP-2              | 4      | 1.17            | 9.97                     | 8.64                       | 84.9    |          |
| L2483-06 | 6FT-TP-3              | 5      | 1.17            | 9.81                     | 8.34                       | 83.0    |          |
| L2483-08 | 6FT-TP-4              | 6      | 1.17            | 9.62                     | 8.45                       | 86.2    |          |
| L2483-10 | 6FT-TP-5              | 7      | 1.17            | 9.98                     | 8.5                        | 83.2    |          |
| L2483-12 | 6FT-TP-6              | 8      | 1.17            | 9.78                     | 8.73                       | 87.8    |          |
| L2483-14 | 6FT-TP-7              | 9      | 1.15            | 9.81                     | 6.39                       | 60.5    |          |
| L2483-16 | 6FT-TP-8              | 10     | 1.17            | 9.94                     | 6.18                       | 57.1    |          |
| L2483-18 | 6FT-TP-1-EPH2         | 11     | 1.16            | 9.62                     | 8.6                        | 87.9    |          |
| L2483-19 | 6FT-TP-2-EPH2         | 12     | 1.17            | 9.97                     | 8.64                       | 84.9    |          |
| L2483-20 | 6FT-TP-3-EPH2         | 13     | 1.17            | 9.81                     | 8.34                       | 83.0    |          |
| L2483-21 | 6FT-TP-4-EPH2         | 14     | 1.17            | 9.62                     | 8.45                       | 86.2    |          |
| L2483-22 | 6FT-TP-5-EPH2         | 15     | 1.17            | 9.98                     | 8.5                        | 83.2    |          |
| L2483-23 | 6FT-TP-6-EPH2         | 16     | 1.17            | 9.78                     | 8.73                       | 87.8    |          |
| L2483-24 | 6FT-TP-7-EPH2         | 17     | 1.15            | 9.81                     | 6.39                       | 60.5    |          |
| L2483-25 | 6FT-TP-8-EPH2         | 18     | 1.17            | 9.94                     | 6.18                       | 57.1    |          |
| L2490-01 | TT-019-IDWSO-20200504 | 19     | 1.16            | 9.7                      | 7.25                       | 71.3    |          |
| L2491-02 | 6FT-TP-9              | 20     | 1.16            | 9.6                      | 6.09                       | 58.4    |          |
| L2491-03 | 6FT-TP-9-EPH-2        | 21     | 1.16            | 9.6                      | 6.09                       | 58.4    |          |
| L2491-05 | 6FT-TP-10             | 22     | 1.17            | 9.98                     | 5.56                       | 49.8    |          |
| L2491-06 | 6FT-TP-10-EPH-2       | 23     | 1.17            | 9.98                     | 5.56                       | 49.8    |          |
| L2491-08 | 6FT-TP-11             | 24     | 1.17            | 9.8                      | 6.61                       | 63.0    |          |
| L2491-09 | 6FT-TP-11-EPH-2       | 25     | 1.17            | 9.8                      | 6.61                       | 63.0    |          |
| L2491-11 | 6FT-TP-12             | 26     | 1.17            | 9.75                     | 9.12                       | 92.7    |          |
| L2491-12 | 6FT-TP-12-EPH-2       | 27     | 1.17            | 9.75                     | 9.12                       | 92.7    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

L5109003

WorkList Name : %1-050520

WorkList ID : 138460

Department : Wet-Chemistry

Date : 05-05-2020 15:18:55

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample       | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------------|--------------|--------------|
| 04/30/2020 | Solid  | L2401-09 | Percent Solids | Cool 4 deg C | PSEG01   | J13              | DRAINAGE-SEWALE       | 04/23/2020   | Chemtech -SO |
| 04/30/2020 | Solid  | L2401-10 | Percent Solids | Cool 4 deg C | PSEG01   | J13              | DRAINAGE-SEWALE       | 04/23/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-02 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-1              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-04 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-2              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-06 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-3              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-08 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-4              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-10 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-5              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-12 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-6              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-14 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-7              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-16 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-8              | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-18 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-1-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-19 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-2-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-20 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-3-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-21 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-4-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-22 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-5-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-23 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-6-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-24 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-7-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/14/2020 | Solid  | L2483-25 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-8-EPH2         | 05/04/2020   | Chemtech -SO |
| 05/07/2020 | Solid  | L2490-01 | Percent Solids | Cool 4 deg C | TETR06   | L11              | TT-019-IDWSO-20200504 | 05/04/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-02 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-9              | 05/05/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-03 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-9-EPH-2        | 05/05/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-05 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-10             | 05/05/2020   | Chemtech -SO |

Date/Time 05/05/2020

Received by: 7K

Relinquished by: AM

Date/Time 05/05/2020

Received by: AM

Relinquished by: DK

# WORKLIST(Hardcopy Internal Chain)

LAB 013

WorkList Name : %1-050520

WorkList ID : 138460

Department : Wet-Chemistry

Date : 05-05-2020 15:18:55

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 05/15/2020 | Solid  | L2491-06 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-10-EPH-2 | 05/05/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-08 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-11       | 05/05/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-09 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-11-EPH-2 | 05/05/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-11 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-12       | 05/05/2020   | Chemtech -SO |
| 05/15/2020 | Solid  | L2491-12 | Percent Solids | Cool 4 deg C | PSEG01   | L41              | 6FT-TP-12-EPH-2 | 05/05/2020   | Chemtech -SO |

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: