



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109019

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/06/2020	10:05
2	CAL2	1	Water	NA	NA	20.2	7.00	05/06/2020	10:07
3	CAL3	1	Water	NA	NA	20.2	10.03	05/06/2020	10:09
4	ICV	1	Water	NA	NA	20.3	7.01	05/06/2020	10:10
5	CCV1	1	Water	NA	NA	20.1	2.02	05/06/2020	10:11
6	L2466-01	1	Solid	20.03	20	20.3	5.54	05/06/2020	10:12
7	L2513-04	1	Solid	20.03	20	19.8	6.15	05/06/2020	10:15
8	L2515-04	1	Solid	20.02	20	19.8	9.21	05/06/2020	10:16
9	L2515-04DUP	1	Solid	20.03	20	19.9	9.22	05/06/2020	10:19
10	CCV2	1	Water	NA	NA	20.1	12.02	05/06/2020	10:20

28109619

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY L2513 WorkList ID: 138484 Department: Wet-Chemistry Date: 05-06-2020 10:00:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2020	Solid	L2466-01	Corrosivity	Cool 4 deg C	PSEG01	L31	DRUM-1-3	05/01/2020	9045D
05/05/2020	Solid	L2513-04	Corrosivity	Cool 4 deg C	PSEG05	L41	BU-03-050520	05/05/2020	9045D
05/05/2020	Solid	L2515-04	Corrosivity	Cool 4 deg C	PSEG01	L42	TP-18A	05/05/2020	9045D

Date/Time 05.06.2020 10:00
Received by: Jn
Relinquished by: Jn

Date/Time 05.06.2020 11:30
Received by: Jn
Relinquished by: Jn