



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109020

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/06/2020	10:05
2	CAL2	1	Water	NA	NA	20.2	7.00	05/06/2020	10:07
3	CAL3	1	Water	NA	NA	20.2	10.03	05/06/2020	10:09
4	ICV	1	Water	NA	NA	20.3	7.01	05/06/2020	10:10
5	CCV1	1	Water	NA	NA	20.2	2.01	05/06/2020	10:26
6	L2492-03	1	Water	NA	NA	12.3	9.01	05/06/2020	10:27
7	L2492-03DUP	1	Water	NA	NA	12.4	9.02	05/06/2020	10:28
8	L2492-04	1	Water	NA	NA	12.2	9.31	05/06/2020	10:30
9	L2492-05	1	Water	NA	NA	13.1	9.21	05/06/2020	10:31
10	L2492-06	1	Water	NA	NA	12.2	9.33	05/06/2020	10:33
11	L2492-07	1	Water	NA	NA	12.3	9.31	05/06/2020	10:35
12	L2492-08	1	Water	NA	NA	12.1	9.32	05/06/2020	10:36
13	L2492-09	1	Water	NA	NA	13.2	9.33	05/06/2020	10:38
14	L2492-10	1	Water	NA	NA	18.1	9.27	05/06/2020	10:40
15	L2492-11	1	Water	NA	NA	13.2	9.28	05/06/2020	10:42
16	CCV2	1	Water	NA	NA	20.1	12.02	05/06/2020	10:45
17	L2492-12	1	Water	NA	NA	20.1	8.02	05/06/2020	10:46
18	CCV3	1	Water	NA	NA	20.2	2.02	05/06/2020	10:48

28109020

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2492W WorkList ID : 138485 Department : Wet-Chemistry Date : 05-06-2020 10:00:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/05/2020	Water	L2492-03	pH	Cool 4 deg C	PSEG01	L42	30981	05/05/2020	9040C
05/05/2020	Water	L2492-04	pH	Cool 4 deg C	PSEG01	L42	30978	05/05/2020	9040C
05/05/2020	Water	L2492-05	pH	Cool 4 deg C	PSEG01	L42	30977	05/05/2020	9040C
05/05/2020	Water	L2492-06	pH	Cool 4 deg C	PSEG01	L42	30976	05/05/2020	9040C
05/05/2020	Water	L2492-07	pH	Cool 4 deg C	PSEG01	L42	30975	05/05/2020	9040C
05/05/2020	Water	L2492-08	pH	Cool 4 deg C	PSEG01	L42	30974	05/05/2020	9040C
05/05/2020	Water	L2492-09	pH	Cool 4 deg C	PSEG01	L42	30973	05/05/2020	9040C
05/05/2020	Water	L2492-10	pH	Cool 4 deg C	PSEG01	L42	30972	05/05/2020	9040C
05/05/2020	Water	L2492-11	pH	Cool 4 deg C	PSEG01	L42	30971	05/05/2020	9040C
05/05/2020	Water	L2492-12	pH	Cool 4 deg C	PSEG01	L42	30970	05/05/2020	9040C

Date/Time 05.06.2020 10:00
Received by: 3n
Relinquished by: cp

Date/Time 05.06.2020 11:30
Received by: CP
Relinquished by: 71