


LB109021

7471B

INSTRUMENT ID: CV1

 μ Abs.:

52454



Linear ▼

A= 0.0000e+000

B= 1.9170e-004

C= 1.7166e-003

Rho= 0.9998772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-2	0.000	-2				
0.2	0.200	0.214	0.014	1107	0.0 %	1107				
2.5	2.500	2.480	-0.020	12929	0.0 %	12929				
5.0	5.000	5.060	0.060	26387	0.0 %	26387				
7.5	7.500	7.388	-0.112	38529	0.0 %	38529				
10.0	10.000	10.057	0.057	52454	0.0 %	52454				

LB109021 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 May 2020 10:22:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7471B	PPB	06 May 2020 10:25:17	S	Std
0.2 - 1	50.2	1107	-	0.2000	7471B	PPB	06 May 2020 10:27:34	S	Std
2.5 - 1	52.5	12929	-	2.5000	7471B	PPB	06 May 2020 10:29:51	S	Std
5.0 - 1	55	26387	-	5.0000	7471B	PPB	06 May 2020 10:32:08	S	Std
7.5 - 1	57.5	38529	-	7.5000	7471B	PPB	06 May 2020 10:34:24	S	Std
10.0 - 1	510	52454	-	10.0000	7471B	PPB	06 May 2020 10:36:41	S	Std
ICV32 - 1	ICV32	19624	3.7636	-7471B	PPB	06 May 2020 10:39:33	U	SMPL	
ICB32 - 1	ICB32	-127	-0.0226	-7471B	PPB	06 May 2020 10:41:48	U	SMPL	
CCV61 - 1	CCV61	26687	5.1175	-7471B	PPB	06 May 2020 10:46:36	U	SMPL	
CCB61 - 1	CCB61	-117	-0.0207	-7471B	PPB	06 May 2020 10:48:51	U	SMPL	
CRA - 1	CRA	1083	0.2093	-7471B	PPB	06 May 2020 10:51:07	U	SMPL	
HighStd - 1	HighStd	52321	10.0315	-7471B	PPB	06 May 2020 10:53:22	U	SMPL	
ChkStd - 1	ChkStd	35934	6.8901	-7471B	PPB	06 May 2020 10:55:37	U	SMPL	
PB128671BL - 1	PBS	-134	-0.0240	-7471B	PPB	06 May 2020 10:57:55	U	SMPL	
PB128671BS - 1	LCSS	19335	3.7082	-7471B	PPB	06 May 2020 11:00:12	U	SMPL	
L2481-11DL2X25 - 1	HSM-04-W	33285	6.3823	-7471B	PPB	06 May 2020 11:02:28	U	SMPL	
L2483-02 - 1	6FT-TP-1	4314	0.8287	-7471B	PPB	06 May 2020 11:04:45	U	SMPL	
L2483-04 - 1	6FT-TP-2	6923	1.3288	-7471B	PPB	06 May 2020 11:07:04	U	SMPL	
L2483-06 - 1	6FT-TP-3	11684	2.2415	-7471B	PPB	06 May 2020 11:09:20	U	SMPL	
L2483-08 - 1	6FT-TP-4	1140	0.2203	-7471B	PPB	06 May 2020 11:11:36	U	SMPL	
CCV62 - 1	CCV62	26365	5.0558	-7471B	PPB	06 May 2020 11:13:50	U	SMPL	
CCB62 - 1	CCB62	-82	-0.0140	-7471B	PPB	06 May 2020 11:16:06	U	SMPL	
L2483-10 - 1	6FT-TP-5	9482	1.8194	-7471B	PPB	06 May 2020 11:18:22	U	SMPL	
L2483-12 - 1	6FT-TP-6	3847	0.7392	-7471B	PPB	06 May 2020 11:20:36	U	SMPL	
L2483-14 - 1	6FT-TP-7	16440	3.1532	-7471B	PPB	06 May 2020 11:22:52	U	SMPL	
L2483-16 - 1	6FT-TP-8	1753	0.3378	-7471B	PPB	06 May 2020 11:25:07	U	SMPL	
L2491-02 - 1	6FT-TP-9	59272	11.3640	-7471B	PPB	06 May 2020 11:27:23	U	SMPL	
L2491-05 - 1	6FT-TP-10	14484	2.7783	-7471B	PPB	06 May 2020 11:29:40	U	SMPL	
L2491-08 - 1	6FT-TP-11	60964	11.6883	-7471B	PPB	06 May 2020 11:31:58	U	SMPL	
L2491-11 - 1	6FT-TP-12	116	0.0240	-7471B	PPB	06 May 2020 11:34:15	U	SMPL	
L2515-02 - 1	TP-18A	7758	1.4889	-7471B	PPB	06 May 2020 11:36:34	U	SMPL	
L2515-02DUP - 1	TP-18ADUP	7291	1.3994	-7471B	PPB	06 May 2020 11:38:50	U	SMPL	
CCV63 - 1	CCV63	26603	5.1014	-7471B	PPB	06 May 2020 11:41:05	U	SMPL	
CCB63 - 1	CCB63	-78	-0.0132	-7471B	PPB	06 May 2020 11:43:20	U	SMPL	
L2515-02MS - 1	TP-18AMS	30000	5.7526	-7471B	PPB	06 May 2020 11:45:36	U	SMPL	
L2515-02MSD - 1	TP-18AMSD	28334	5.4332	-7471B	PPB	06 May 2020 11:47:52	U	SMPL	
L2515-02LX5 - 1		1296	0.2502	-7471B	PPB	06 May 2020 11:50:08	U	SMPL	
L2515-02A - 1		28904	5.5425	-7471B	PPB	06 May 2020 11:52:25	U	SMPL	
L2491-02DLX2 - 1		29104	5.5809	-7471B	PPB	06 May 2020 11:54:41	U	SMPL	
L2491-08DLX2 - 1		29170	5.5935	-7471B	PPB	06 May 2020 11:56:58	U	SMPL	
CCV64 - 1	CCV64	26598	5.1005	-7471B	PPB	06 May 2020 11:59:16	U	SMPL	
CCB64 - 1	CCB64	-89	-0.0153	-7471B	PPB	06 May 2020 12:01:34	U	SMPL	