



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109022

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/06/2020	10:05
2	CAL2	1	Water	NA	NA	20.2	7.00	05/06/2020	10:06
3	CAL3	1	Water	NA	NA	20.2	10.03	05/06/2020	10:09
4	ICV	1	Water	NA	NA	20.3	7.01	05/06/2020	10:10
5	CCV1	1	Water	NA	NA	20.2	2.01	05/06/2020	10:27
6	L2492-15	1	Solid	20.03	20	19.9	6.25	05/06/2020	10:50
7	L2492-16	1	Solid	20.02	20	19.7	6.80	05/06/2020	11:00
8	L2492-17	1	Solid	20.03	20	18.6	6.90	05/06/2020	11:01
9	L2492-17DUP	1	Solid	20.04	20	18.7	6.91	05/06/2020	11:02
10	CCV2	1	Water	NA	NA	20.1	12.02	05/06/2020	11:05

187109022

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L2492 WorkList ID : 138486 Department : Wet-Chemistry Date : 05-06-2020 10:02:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/05/2020	Solid	L2492-15	pH	Cool 4 deg C	PSEG01	L42	30910	05/05/2020	9045D
05/05/2020	Solid	L2492-16	pH	Cool 4 deg C	PSEG01	L42	30980	05/05/2020	9045D
05/05/2020	Solid	L2492-17	pH	Cool 4 deg C	PSEG01	L42	30979	05/05/2020	9045D

Date/Time 05-06-2020 10:10
 Received by: JD
 Relinquished by: JD

Date/Time 05-06-2020 10:35
 Received by: JD
 Relinquished by: JD