


LB109025

ISM02.4

INSTRUMENT ID: CV1

 μ Abs.:

44904



Linear

A= 0.0000e+000

B= 2.2472e-004

C= 6.8110e-003

Rho= 0.9998494

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/D
0.00	0.000	0.028	0.028	93	0.000	93					
0.05	0.050					279					
0.20	0.200	0.213	0.013	918	0.0 %	918					
2.50	2.500	2.515	0.015	11161	0.0 %	11161					
5.00	5.000	4.956	-0.044	22025	0.0 %	22025					
7.50	7.500	7.390	-0.110	32857	0.0 %	32857					
10.0	10.000	10.098	0.098	44904	0.0 %	44904					

0/D
 -
 67 NB
 511th

LB109025
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	93	-		0 ISM02.4	PPB	5/6/2020 12:25	Std
	0.2 S0.2	918	-		0.2 ISM02.4	PPB	5/6/2020 12:27	Std
	2.5 S2.5	11161	-		2.5 ISM02.4	PPB	5/6/2020 12:30	Std
	5 S5.0	22025	-		5 ISM02.4	PPB	5/6/2020 12:32	Std
	7.5 S7.5	32857	-		7.5 ISM02.4	PPB	5/6/2020 12:34	Std
	10 S10.0	44904	-		10 ISM02.4	PPB	5/6/2020 12:37	Std
ICV	ICV	17222	3.8769 -		ISM02.4	PPB	5/6/2020 12:39	SMPL
ICB	ICB	-50	-0.0044 -		ISM02.4	PPB	5/6/2020 12:42	SMPL
CCV065	CCV065	21813	4.9086 -		ISM02.4	PPB	5/6/2020 12:44	SMPL
CCB065	CCB065	-68	-0.0085 -		ISM02.4	PPB	5/6/2020 12:46	SMPL
PB128681BL	PBW681	13	0.0097 -		ISM02.4	PPB	5/6/2020 12:48	SMPL
L2489-01	MBFSM8	89	0.0268 -		ISM02.4	PPB	5/6/2020 12:51	SMPL
L2489-02	MBFSM8D	78	0.0243 -		ISM02.4	PPB	5/6/2020 12:53	SMPL
L2489-03	MBFSM8S	4614	1.0437 -		ISM02.4	PPB	5/6/2020 12:55	SMPL
L2489-04	MBFSM9	-9	0.0048 -		ISM02.4	PPB	5/6/2020 12:57	SMPL
L2489-05	MBFSN0	274	0.0684 -		ISM02.4	PPB	5/6/2020 13:00	SMPL
L2489-06	MBFSN2	175	0.0461 -		ISM02.4	PPB	5/6/2020 13:02	SMPL
CCV066	CCV066	21937	4.9365 -		ISM02.4	PPB	5/6/2020 13:04	SMPL
CCB066	CCB066	-51	-0.0046 -		ISM02.4	PPB	5/6/2020 13:06	SMPL