



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109027

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/06/2020	13:10
2	CAL2	1	Water	NA	NA	20.1	7.01	05/06/2020	13:11
3	CAL3	1	Water	NA	NA	20.1	10.03	05/06/2020	13:12
4	ICV	1	Water	NA	NA	20.2	6.99	05/06/2020	13:15
5	CCV1	1	Water	NA	NA	20.2	2.01	05/06/2020	13:16
6	L2119-01	1	Water	NA	NA	18.7	5.66	05/06/2020	13:19
7	L2121-01	1	Water	NA	NA	18.6	5.20	05/06/2020	13:20
8	L2121-01DUP	1	Water	NA	NA	18.7	5.21	05/06/2020	13:21
9	CCV2	1	Water	NA	NA	20.2	12.02	05/06/2020	13:22

18109027

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2521

WorkList ID : 138496

Department : Wet-Chemistry

Date : 05-06-2020 12:53:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/05/2020	Water	L2119-01-17	pH	Cool 4 deg C	STER04	L41	POOL-WATER	05/05/2020	SM 4500-H B
05/05/2020	Water	L2121-01-17	pH	Cool 4 deg C	STER05	L41	BOTTLE-1	05/05/2020	SM 4500-H B