



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109034

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/06/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	05/06/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.04	05/06/2020	17:19
4	ICV	1	Water	NA	NA	20.2	6.98	05/06/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.02	05/06/2020	17:22
6	L2114-01	1	Solid	20.03	20	20.4	8.17	05/06/2020	17:23
7	L2497-04	1	Solid	20.04	20	20.4	7.52	05/06/2020	17:25
8	L2497-04DUP	1	Solid	20.03	20	20.5	7.53	05/06/2020	17:26
9	L2498-08	1	Solid	20.02	20	20.5	8.31	05/06/2020	17:29
10	L2523-04	1	Solid	20.01	20	20.3	9.31	05/06/2020	17:30
11	L2525-03	1	Solid	20.03	20	20.3	8.09	05/06/2020	17:32
12	L2530-02	1	Solid	20.05	20	20.1	7.30	05/06/2020	17:33
13	CCV2	1	Water	NA	NA	20.1	12.02	05/06/2020	17:35

WORKLIST(Hardcopy Internal Chain)

109034

WorkList Name : CORROSIVITY L2523 WorkList ID : 138503 Department : Wet-Chemistry Date : 05-06-2020 14:55:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/06/2020	Solid	L2114-01	Corrosivity	Cool 4 deg C	PSEG01	L42	PEN-DRUM	05/06/2020	9045D
05/06/2020	Solid	L2497-04	Corrosivity	Cool 4 deg C	PSEG05	L51	HD-01-050620	05/06/2020	9045D
05/06/2020	Solid	L2498-08	Corrosivity	Cool 4 deg C	PSEG05	L51	CL-02-050620	05/06/2020	9045D
05/06/2020	Solid	L2523-04	Corrosivity	Cool 4 deg C	PSEG01	L43	HDD	05/06/2020	9045D
05/06/2020	Solid	L2525-03	Corrosivity	Cool 4 deg C	PSEG01	L43	DRUM-1-3	05/06/2020	9045D
05/06/2020	Solid	L2530-02	Corrosivity	Cool 4 deg C	PSEG01	L51	RO-14	05/06/2020	9045D

Date/Time 05.06.2020 16:00
 Received by: Te
 Relinquished by: Jo

Date/Time 05.06.2020 18:00
 Received by: 70
 Relinquished by: 30