


LB109037

7470A

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	-15	0.000	-15				
0.2	0.200	0.210	0.010	708	0.0 %	708				
2.5	2.500	2.433	-0.067	9613	0.0 %	9613				
5.0	5.000	4.960	-0.040	19741	0.0 %	19741				
7.5	7.500	7.635	0.135	30458	0.0 %	30458				
10.0	10.000	9.932	-0.068	39662	0.0 %	39662				

LB109037 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 May 2020 08:44:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-15	-	0.0000	7470A	PPB	07 May 2020 08:48:38	S	Std
0.2 - 1	50.2	708	-	0.2000	7470A	PPB	07 May 2020 08:50:46	S	Std
2.5 - 1	52.5	9613	-	2.5000	7470A	PPB	07 May 2020 08:52:54	S	Std
5.0 - 1	55	19741	-	5.0000	7470A	PPB	07 May 2020 08:55:01	S	Std
7.5 - 1	57.5	30458	-	7.5000	7470A	PPB	07 May 2020 08:57:09	S	Std
10.0 - 1	59	39662	-	10.0000	7470A	PPB	07 May 2020 08:59:15	S	Std
ICV34 - 1	ICV34	15703	3.9525	-	7470A	PPB	07 May 2020 09:01:55	U	SMPL
ICB34 - 1	ICB34	3	0.0342	-	7470A	PPB	07 May 2020 09:06:24	U	SMPL
CCV75 - 1	CCV75	19382	4.8707	-	7470A	PPB	07 May 2020 09:08:29	U	SMPL
CCB75 - 1	CCB75	-83	0.0127	-	7470A	PPB	07 May 2020 09:10:34	U	SMPL
CRA - 1	CRA	702	0.2086	-	7470A	PPB	07 May 2020 09:12:40	U	SMPL
HighStd - 1	HighStd	40355	10.1051	-	7470A	PPB	07 May 2020 09:14:46	U	SMPL
ChkStd - 1	ChkStd	27177	6.8162	-	7470A	PPB	07 May 2020 09:16:52	U	SMPL
PB128695BL - 1	PBW	-107	0.0067	-	7470A	PPB	07 May 2020 09:19:00	U	SMPL
PB128695BS - 1	LCSW	14640	3.6872	-	7470A	PPB	07 May 2020 09:21:07	U	SMPL
L2464-01 - 1	TRE-50006	-6	0.0319	-	7470A	PPB	07 May 2020 09:23:14	U	SMPL
L2483-01 - 1	6FT-TP-1	76	0.0524	-	7470A	PPB	07 May 2020 09:25:21	U	SMPL
L2483-03 - 1	6FT-TP-2	126	0.0649	-	7470A	PPB	07 May 2020 09:27:29	U	SMPL
L2483-05 - 1	6FT-TP-3	105	0.0596	-	7470A	PPB	07 May 2020 09:29:36	U	SMPL
L2483-07 - 1	6FT-TP-4	103	0.0591	-	7470A	PPB	07 May 2020 09:31:43	U	SMPL
CCV76 - 1	CCV76	19549	4.9124	-	7470A	PPB	07 May 2020 09:33:48	U	SMPL
CCB76 - 1	CCB76	-91	0.0107	-	7470A	PPB	07 May 2020 09:35:54	U	SMPL
L2483-09 - 1	6FT-TP-5	146	0.0699	-	7470A	PPB	07 May 2020 09:37:59	U	SMPL
L2483-11 - 1	6FT-TP-6	143	0.0691	-	7470A	PPB	07 May 2020 09:40:05	U	SMPL
L2483-13 - 1	6FT-TP-7	160	0.0734	-	7470A	PPB	07 May 2020 09:42:11	U	SMPL
L2483-15 - 1	6FT-TP-8	156	0.0724	-	7470A	PPB	07 May 2020 09:44:17	U	SMPL
L2491-01 - 1	6FT-TP-9	148	0.0704	-	7470A	PPB	07 May 2020 09:46:24	U	SMPL
L2491-04 - 1	6FT-TP-10	137	0.0676	-	7470A	PPB	07 May 2020 09:48:31	U	SMPL
L2491-07 - 1	6FT-TP-11	114	0.0619	-	7470A	PPB	07 May 2020 09:50:38	U	SMPL
L2491-10 - 1	6FT-TP-12	103	0.0591	-	7470A	PPB	07 May 2020 09:52:46	U	SMPL
L2492-13 - 1	OILY-DEBRIS-COMP	188	0.0803	-	7470A	PPB	07 May 2020 09:54:54	U	SMPL
L2513-04 - 1	BU-03-050520	90	0.0559	-	7470A	PPB	07 May 2020 09:57:00	U	SMPL
CCV77 - 1	CCV77	19291	4.8480	-	7470A	PPB	07 May 2020 09:59:06	U	SMPL
CCB77 - 1	CCB77	-125	0.0022	-	7470A	PPB	07 May 2020 10:01:12	U	SMPL
L2515-04 - 1	TP-18A	-3	0.0327	-	7470A	PPB	07 May 2020 10:03:18	U	SMPL
L2515-04DUP - 1	TP-18ADUP	85	0.0546	-	7470A	PPB	07 May 2020 10:05:24	U	SMPL
L2515-04MS - 1	TP-18AMS	14208	3.5794	-	7470A	PPB	07 May 2020 10:07:30	U	SMPL
L2515-04MSD - 1	TP-18AMSD	14875	3.7459	-	7470A	PPB	07 May 2020 10:09:37	U	SMPL
L2515-04LX5 - 1		-84	0.0125	-	7470A	PPB	07 May 2020 10:11:44	U	SMPL
L2515-04A - 1		15023	3.7828	-	7470A	PPB	07 May 2020 10:13:51	U	SMPL
PB128695TB - 1	PB128667TB	20	0.0384	-	7470A	PPB	07 May 2020 10:15:57	U	SMPL
CCV78 - 1	CCV78	19318	4.8548	-	7470A	PPB	07 May 2020 10:18:05	U	SMPL
CCB78 - 1	CCB78	-67	0.0167	-	7470A	PPB	07 May 2020 10:20:13	U	SMPL