



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB109051

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP84092
Turbidity Calibration std, 0NTU	WP84111
Turbidity Calibration std, 20NTU	WP84088
Turbidity Calibration std, 40NTU	WP84087
Turbidity Calibration std, 1NTU	WP84091
Turbidity Calibration std, 5NTU	WP84090
Turbidity Calibration - CCV std, 10 NTU	WP84089
10 NTU Standard 500 ml	W2489

Intercept: 0.1414

Slope: 1.12

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.266	05/07/2020	09:30
2	CAL2	1	1	1.300	05/07/2020	09:34
3	CAL3	5	1	5.900	05/07/2020	09:38
4	CAL4	10	1	11.100	05/07/2020	09:42
5	CAL5	20	1	22.300	05/07/2020	09:46
6	CAL6	40	1	45.100	05/07/2020	09:50



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB109051

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.600	9.338	05/07/2020	09:55
2	ICB		1	0.262	0.108	05/07/2020	10:02
3	CCV1	10	1	10.900	9.606	05/07/2020	10:06
4	CCB1		1	0.277	0.121	05/07/2020	10:10
5	LB109051BL		1	0.273	0.118	05/07/2020	10:15
6	L2182-12		1	0.748	0.542	05/07/2020	10:19
7	CCV2	10	1	10.800	9.517	05/07/2020	10:24
8	CCB2		1	0.285	0.128	05/07/2020	10:28