


LB109052

ISM02.4

INSTRUMENT 101 CV1

 μ Abs.:

4347E



Linear

A= 0.0000e+000

B= 2.3154e-004

C= -6.8834e-003

Rho= 0.9999211

Accept=Accepted

slope
y = 1.215e-004

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.004	0.004	46	0.000	46					1
0.05	0.050					279					3
0.20	0.200	0.205	0.005	913	0.0 %	913					2
2.50	2.500	2.549	0.049	11038	0.0 %	11038					1
5.00	5.000	4.943	-0.057	21378	0.0 %	21378					1
7.50	7.500	7.440	-0.060	32162	0.0 %	32162					1
10.0	10.000	10.060	0.060	43478	0.0 %	43478					1

10D
1
3
2
1
1
1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	46	-		0 ISM02.4	PPB	5/7/2020 13:24	Std
	0.2 S0.2	913	-		0.2 ISM02.4	PPB	5/7/2020 13:26	Std
	2.5 S2.5	11038	-		2.5 ISM02.4	PPB	5/7/2020 13:28	Std
	5 S5.0	21378	-		5 ISM02.4	PPB	5/7/2020 13:31	Std
	7.5 S7.5	32162	-		7.5 ISM02.4	PPB	5/7/2020 13:33	Std
	10 S10.0	43478	-		10 ISM02.4	PPB	5/7/2020 13:38	Std
ICV	ICV	17352	4.0108 -		ISM02.4	PPB	5/7/2020 13:40	SMPL
ICB	ICB	-131	-0.0372 -		ISM02.4	PPB	5/7/2020 13:43	SMPL
CCV082	CCV082	21428	4.9545 -		ISM02.4	PPB	5/7/2020 13:45	SMPL
CCB082	CCB082	-127	-0.0363 -		ISM02.4	PPB	5/7/2020 13:47	SMPL
PB128710BL	PBW710	81	0.0119 -		ISM02.4	PPB	5/7/2020 13:49	SMPL
L2518-01	MBFSN3	192	0.0376 -		ISM02.4	PPB	5/7/2020 13:52	SMPL
L2518-02	MBFSN6	171	0.0327 -		ISM02.4	PPB	5/7/2020 13:54	SMPL
L2518-03	MBFSN6D	187	0.0364 -		ISM02.4	PPB	5/7/2020 13:56	SMPL
L2518-04	MBFSN6S	4747	1.0922 -		ISM02.4	PPB	5/7/2020 13:58	SMPL
L2518-05	MBFSP0	192	0.0376 -		ISM02.4	PPB	5/7/2020 14:01	SMPL
CCV083	CCV083	21417	4.952 -		ISM02.4	PPB	5/7/2020 14:03	SMPL
CCB083	CCB083	-104	-0.031 -		ISM02.4	PPB	5/7/2020 14:05	SMPL