



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109055

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/07/2020	14:00
2	CAL2	1	Water	NA	NA	20.2	7.01	05/07/2020	14:01
3	CAL3	1	Water	NA	NA	20.2	10.03	05/07/2020	14:02
4	ICV	1	Water	NA	NA	20.2	7.00	05/07/2020	14:05
5	CCV1	1	Water	NA	NA	20.1	2.01	05/07/2020	14:06
6	L2535-01	1	Solid	20.03	20	19.4	8.61	05/07/2020	14:10
7	L2535-01DUP	1	Solid	20.04	20	19.5	8.62	05/07/2020	14:11
8	L2535-03	1	Solid	20.02	20	19.5	8.70	05/07/2020	14:15
9	CCV2	1	Water	NA	NA	20.2	12.02	05/07/2020	14:16

LB109055

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L2535 WorkList ID : 138548 Department : Wet-Chemistry Date : 05-07-2020 13:46:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2020	Solid	L2535-01	pH	Cool 4 deg C	PSEG01	L53	COMP-02-06	05/07/2020	9045D
05/07/2020	Solid	L2535-03	pH	Cool 4 deg C	PSEG01	L53	COMP-07-09	05/07/2020	9045D

Date/Time 05/07/2020 13:50
Received by: Jg
Relinquished by: Jg

Date/Time 05/07/2020 15:10
Received by: Jg
Relinquished by: Jg