

LB109058


7470A

INSTRUMENT ID: CY 1

Linear

 μ Abs.:

39229



A= 0.0000e+000

B= 2.5501e-004

C= -4.7956e-002

Rho= 0.9999681

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	113	0.000	113				
0.2	0.200	0.179	-0.021	891	0.0 %	891				
2.5	2.500	2.521	0.021	10073	0.0 %	10073				
5.0	5.000	5.034	0.034	19930	0.0 %	19930				
7.5	7.500	7.529	0.029	29713	0.0 %	29713				
10.0	10.000	9.956	-0.044	39229	0.0 %	39229				

LB109058 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 May 2020 15:13:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	113	-	0.0000	7470A	PPB	07 May 2020 15:24:49	S	Std
0.2 - 1	SP.2	891	-	0.2000	7470A	PPB	07 May 2020 15:26:57	S	Std
2.5 - 1	SP.5	10073	-	2.5000	7470A	PPB	07 May 2020 15:29:04	S	Std
5.0 - 1	SP.5	19930	-	5.0000	7470A	PPB	07 May 2020 15:31:12	S	Std
7.5 - 1	SP.5	29713	-	7.5000	7470A	PPB	07 May 2020 15:33:19	S	Std
10.0 - 1	SP.1	39229	-	10.0000	7470A	PPB	07 May 2020 15:35:25	S	Std
ICV36 - 1	ICV36	15192	3.8261	-	7470A	PPB	07 May 2020 15:38:08	U	SMPL
ICB36 - 1	ICB36	-36	-0.0571	-	7470A	PPB	07 May 2020 15:40:14	U	SMPL
CCV84 - 1	CCV84	19890	5.0241	-	7470A	PPB	07 May 2020 15:42:20	U	SMPL
CCB84 - 1	CCB84	-42	-0.0587	-	7470A	PPB	07 May 2020 15:44:25	U	SMPL
CRA - 1	CRA	845	0.1675	-	7470A	PPB	07 May 2020 15:46:31	U	SMPL
HighStd - 1	HighStd	39341	9.9843	-	7470A	PPB	07 May 2020 15:48:37	U	SMPL
ChkStd - 1	ChkStd	27304	6.9148	-	7470A	PPB	07 May 2020 15:50:43	U	SMPL
PB128719BL - 1	PBW	-45	-0.0594	-	7470A	PPB	07 May 2020 15:52:51	U	SMPL
PB128719BS - 1	LCSW	16495	4.1584	-	7470A	PPB	07 May 2020 15:54:58	U	SMPL
L2497-04 - 1	HD-01-050620	-11	-0.0508	-	7470A	PPB	07 May 2020 15:57:05	U	SMPL
L2497-04DUP - 1	HD-01-050620DUP	54	-0.0342	-	7470A	PPB	07 May 2020 15:59:13	U	SMPL
L2497-04MS - 1	HD-01-050620MS	13967	3.5137	-	7470A	PPB	07 May 2020 16:01:20	U	SMPL
L2497-04MSD - 1	HD-01-050620MSD	14361	3.6142	-	7470A	PPB	07 May 2020 16:03:29	U	SMPL
L2498-08 - 1	CL-02-050620	88	-0.0255	-	7470A	PPB	07 May 2020 16:05:35	U	SMPL
CCV85 - 1	CCV85	20157	5.0922	-	7470A	PPB	07 May 2020 16:07:41	U	SMPL
CCB85 - 1	CCB85	-50	-0.0607	-	7470A	PPB	07 May 2020 16:09:47	U	SMPL
L2523-04 - 1	HDD	250	0.0158	-	7470A	PPB	07 May 2020 16:11:52	U	SMPL
L2525-03 - 1	DRUM-1-3	370	0.0464	-	7470A	PPB	07 May 2020 16:13:58	U	SMPL
L2530-02 - 1	RO-14	262	0.0189	-	7470A	PPB	07 May 2020 16:16:04	U	SMPL
PB128720BL - 1	PBW	142	-0.0117	-	7470A	PPB	07 May 2020 16:18:11	U	SMPL
L2113-01 - 1	COMP	201	0.0033	-	7470A	PPB	07 May 2020 16:22:24	U	SMPL
L2113-01DUP - 1	COMPDUP	196	0.0020	-	7470A	PPB	07 May 2020 16:24:32	U	SMPL
L2113-01MS - 1	COMPMS	16716	4.2147	-	7470A	PPB	07 May 2020 16:29:06	U	SMPL
L2113-01MSD - 1	COMPMSD	16667	4.2023	-	7470A	PPB	07 May 2020 16:31:14	U	SMPL
CCV86 - 1	CCV86	19286	4.8701	-	7470A	PPB	07 May 2020 16:33:20	U	SMPL
CCB86 - 1	CCB86	-15	-0.0518	-	7470A	PPB	07 May 2020 16:35:26	U	SMPL
L2497-04LX5 - 1		73	-0.0293	-	7470A	PPB	07 May 2020 16:37:33	U	SMPL
L2497-04A - 1		15383	3.8748	-	7470A	PPB	07 May 2020 16:39:40	U	SMPL
L2113-01LX5 - 1		-13	-0.0513	-	7470A	PPB	07 May 2020 16:41:46	U	SMPL
L2113-01A - 1		16420	4.1393	-	7470A	PPB	07 May 2020 16:43:52	U	SMPL
PB128719TB - 1	PB128696TB	24	-0.0418	-	7470A	PPB	07 May 2020 16:45:58	U	SMPL
PB128720BS - 1	LCSW	14938	3.7613	-	7470A	PPB	07 May 2020 16:48:05	U	SMPL
CCV87 - 1	CCV87	19313	4.8770	-	7470A	PPB	07 May 2020 16:50:12	U	SMPL
CCB87 - 1	CCB87	-68	-0.0653	-	7470A	PPB	07 May 2020 16:52:19	U	SMPL