



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109061

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/07/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.01	05/07/2020	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	05/07/2020	17:22
4	ICV	1	Water	NA	NA	20.2	7.00	05/07/2020	17:23
5	CCV1	1	Water	NA	NA	20.2	2.01	05/07/2020	17:25
6	L2495-05	1	Solid	20.03	20	20.3	9.95	05/07/2020	17:26
7	L2503-05	1	Solid	20.02	20	20.4	8.97	05/07/2020	17:29
8	L2506-05	1	Solid	20.03	20	20.7	9.36	05/07/2020	17:30
9	L2531-04	1	Solid	20.02	20	19.6	6.61	05/07/2020	17:32
10	L2531-08	1	Solid	20.03	20	19.5	8.87	05/07/2020	17:33
11	L2554-02	1	Solid	20.02	20	20.3	8.12	05/07/2020	17:35
12	L2554-04	1	Solid	20.03	20	20.3	8.91	05/07/2020	17:36
13	L2557-04	1	Solid	20.02	20	19.9	8.32	05/07/2020	17:39
14	L2557-08	1	Solid	20.03	20	19.8	9.93	05/07/2020	17:40
15	L2557-08DUP	1	Solid	20.04	20	19.9	9.94	05/07/2020	17:41
16	CCV2	1	Water	NA	NA	20.2	12.02	05/07/2020	17:43

WORKLIST(Hardcopy Internal Chain)

VB 109067

WorkList Name : CORROSIONITY L2531

WorkList ID : 138545

Department : Wet-Chemistry

Date : 05-07-2020 13:45:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2020	Solid	L2495-05	Corrosivity	Cool 4 deg C	PSEG05	L61	HR-05-050720	05/07/2020	9045D
05/07/2020	Solid	L2503-05	Corrosivity	Cool 4 deg C	PSEG05	M11	OR-03-050720	05/07/2020	9045D
05/07/2020	Solid	L2506-05	Corrosivity	Cool 4 deg C	PSEG05	L61	TR-04-050720	05/07/2020	9045D
05/06/2020	Solid	L2531-04	Corrosivity	Cool 4 deg C	PSEG01	L52	TP-14A	05/06/2020	9045D
05/06/2020	Solid	L2531-08	Corrosivity	Cool 4 deg C	PSEG01	L52	TP-15A	05/06/2020	9045D
05/07/2020	Solid	L2554-02	Corrosivity	Cool 4 deg C	PSEG01	L62	DRUM-1-7	05/07/2020	9045D
05/07/2020	Solid	L2554-04	Corrosivity	Cool 4 deg C	PSEG01	L62	DRUM-1-1	05/07/2020	9045D
05/07/2020	Solid	L2557-04	Corrosivity	Cool 4 deg C	PSEG01	M12	TP-12A	05/07/2020	9045D
05/07/2020	Solid	L2557-08	Corrosivity	Cool 4 deg C	PSEG01	M12	TP-13A	05/07/2020	9045D

Date/Time 05/07/2020 16:00
 Received by: JS
 Relinquished by: JS

Date/Time 05/07/2020 18:10
 Received by: KS
 Relinquished by: JS