



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109062

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/07/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.01	05/07/2020	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	05/07/2020	17:22
4	ICV	1	Water	NA	NA	20.2	7.00	05/07/2020	17:23
5	CCV1	1	Water	NA	NA	20.2	2.01	05/07/2020	17:45
6	L2549-01	1	Water	NA	NA	21.0	8.83	05/07/2020	17:46
7	L2549-02	1	Water	NA	NA	19.5	5.67	05/07/2020	17:50
8	L2549-03	1	Water	NA	NA	19.8	11.80	05/07/2020	18:00
9	L2549-04	1	Water	NA	NA	19.5	11.68	05/07/2020	18:01
10	L2549-05	1	Water	NA	NA	19.8	11.82	05/07/2020	18:02
11	L2549-05DUP	1	Water	NA	NA	19.9	11.83	05/07/2020	18:03
12	CCV2	1	Water	NA	NA	20.1	12.02	05/07/2020	18:05

WORKLIST(Hardcopy Internal Chain)

28 109062

WorkList Name : ph 12549

WorkList ID : 138553

Department : Wet-Chemistry

Date : 05-07-2020 16:36:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2020	Water	L2549-01 ✓	pH	Cool 4 deg C	PSEG01	L61	VB16173	05/07/2020	9040C
05/07/2020	Water	L2549-02 ✓	pH	Cool 4 deg C	PSEG01	L61	V23626	05/07/2020	9040C
05/07/2020	Water	L2549-03 ✓	pH	Cool 4 deg C	PSEG01	L61	VB25984	05/07/2020	9040C
05/07/2020	Water	L2549-04 ✓	pH	Cool 4 deg C	PSEG01	L61	VB16122	05/07/2020	9040C
05/07/2020	Water	L2549-05 ✓	pH	Cool 4 deg C	PSEG01	L61	VB1524	05/07/2020	9040C

Date/Time 05/07/2020 16:40

Received by: Jn

Relinquished by: KS

Date/Time 05.07.2020 18:00

Received by: JP

Relinquished by: KS