

Analytical Summary Report

Analysis Method: 7196A

ANALYST: Daniel

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB109068

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
Calibration Std. hexchrome 0.1 ppm	WP84138
Calibration Std. hexchrome 0.05 ppm	WP84139
calibration std. hexchrome 0.01 ppm	WP84140
calibration std. hexchrome 0 ppm	WP84141
hexavalent chromium color reagent	WP84005
5N sulfuric acid	WP84066
HNO3 Hex-Chrome, 5M	WP84067
Hexchrome Cleaning Solution	WP84106
Calibration Std Hexachrome 0.025 ppm	WP84143
Hexavalent Chromium ICV-LCS Std	WP84142
Calibration and CCV std HexChrome 0.5PPM	WP84137
Calibration std HexChrome 1.0PPM	WP84136

Intercept: -0.0008

Slope: 0.7972

Regression: 0.999995

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Anal Date	Anal Time
								Backgrnd	Color			
1	CAL1	0	1	100	100		2.05	0.000	0.000	0.000	05/08/2020	12:05
2	CAL2	0.01	1	100	100		2.04	0.000	0.008	0.008	05/08/2020	12:06
3	CAL3	0.025	1	100	100		2.06	0.000	0.020	0.020	05/08/2020	12:07
4	CAL4	0.05	1	100	100		2.03	0.000	0.038	0.038	05/08/2020	12:08
5	CAL5	0.1	1	100	100		2.07	0.000	0.078	0.078	05/08/2020	12:09
6	CAL6	0.5	1	100	100		2.02	0.000	0.397	0.397	05/08/2020	12:10
7	CAL7	1	1	100	100		2.07	0.000	0.797	0.797	05/08/2020	12:11

Analysis Method: 7196A

ANALYST: Daniel

Parameter: Hexavalent Chromiu

SUPERVISOR REVIEW BY: apatel

Run Number: LB109068

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.06	0.000	0.380	0.380	0.478	05/08/2020	12:12
2	ICB		1	100	100		2.04	0.000	0.000	0.000	0.001	05/08/2020	12:13
3	CCV1	0.5	1	100	100		2.08	0.000	0.378	0.378	0.475	05/08/2020	12:14
4	CCB1		1	100	100		2.09	0.000	0.000	0.000	0.001	05/08/2020	12:15
5	PB128748BL		1	2.53	100	7.05	2.04	0.000	0.000	0.000	0.001	05/08/2020	12:16
6	PB128748BS	20	1	2.51	100	7.04	2.06	0.000	0.354	0.354	0.445	05/08/2020	12:17
7	L2495-01		1	2.53	100	7.06	2.04	0.024	0.026	0.002	0.004	05/08/2020	12:18
8	L2495-01DUP		1	2.52	100	7.03	2.03	0.025	0.026	0.001	0.002	05/08/2020	12:19
9	L2495-01MSPre	40	2	2.51	100	7.08	2.07	0.018	0.325	0.307	0.386	05/08/2020	12:20
10	L2495-01MS2Ins	1284	40	2.53	100	7.09	2.09	0.017	0.655	0.638	0.801	05/08/2020	12:21
11	L2495-01MS3Post	40	2	2.50	100	7.07	2.04	0.018	0.354	0.336	0.422	05/08/2020	12:22
12	L2503-01		1	2.51	100	7.04	2.06	0.010	0.011	0.001	0.002	05/08/2020	12:23
13	L2506-01		1	2.50	100	7.03	2.05	0.020	0.022	0.002	0.004	05/08/2020	12:24
14	L2531-02		1	2.52	100	7.05	2.07	0.009	0.010	0.001	0.002	05/08/2020	12:25
15	CCV2	0.5	1	100	100		2.09	0.000	0.384	0.384	0.483	05/08/2020	12:26
16	CCB2		1	100	100		2.07	0.000	0.000	0.000	0.001	05/08/2020	12:27
17	L2531-06		1	2.53	100	7.06	2.04	0.010	0.010	0.000	0.001	05/08/2020	12:28
18	L2554-01		1	2.50	100	7.08	2.03	0.034	0.036	0.002	0.004	05/08/2020	12:29
19	L2554-03		1	2.51	100	7.06	2.01	0.004	0.005	0.001	0.002	05/08/2020	12:30
20	L2557-02		1	2.55	100	7.04	2.04	0.033	0.035	0.002	0.004	05/08/2020	12:31
21	L2557-06		1	2.50	100	7.03	2.09	0.013	0.014	0.001	0.002	05/08/2020	12:32
22	CCV3	0.5	1	100	100		2.07	0.000	0.388	0.388	0.488	05/08/2020	12:33
23	CCB3		1	100	100		2.04	0.000	0.000	0.000	0.001	05/08/2020	12:34