

LB109074


ISM02.4

INSTRUMENT ID: CV1

Linear

μ Abs.:

43993



A= 0.0000e+000

B= 2.2857e-004

C= -1.0819e-002

Rho= 0.9998578

Accept=Accepted

Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	89	0.000	89					1
0.05	0.050	0.040	-0.010	223	0.0 %	223					20
0.20	0.200	0.194	-0.006	897	0.0 %	897					-3
2.50	2.500	2.488	-0.012	10931	0.0 %	10931					0
5.00	5.000	5.098	0.098	22351	0.0 %	22351					2
7.50	7.500	7.376	-0.124	32316	0.0 %	32316					-2
10.0	10.000	10.045	0.045	43993	0.0 %	43993					0

LB109074

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	89	-	0	ISM02.4	PPB	5/8/2020 13:52	Std
	0.05 S0.05	223	-	0.05	ISM02.4	PPB	5/8/2020 13:54	Std
	0.2 S0.2	897	-	0.2	ISM02.4	PPB	5/8/2020 13:56	Std
	2.5 S2.5	10931	-	2.5	ISM02.4	PPB	5/8/2020 13:59	Std
	5 S5.0	22351	-	5	ISM02.4	PPB	5/8/2020 14:01	Std
	7.5 S7.5	32316	-	7.5	ISM02.4	PPB	5/8/2020 14:03	Std
	10 S10.0	43993	-	10	ISM02.4	PPB	5/8/2020 14:06	Std
ICV	ICV	16799	3.829 -		ISM02.4	PPB	5/8/2020 14:08	SMPL
ICB	ICB	85	0.0086 -		ISM02.4	PPB	5/8/2020 14:10	SMPL
CCV093	CCV093	22033	5.0253 -		ISM02.4	PPB	5/8/2020 14:13	SMPL
CCB093	CCB093	66	0.0043 -		ISM02.4	PPB	5/8/2020 14:15	SMPL
PB128742BL	PBW742	147	0.0228 -		ISM02.4	PPB	5/8/2020 14:17	SMPL
L2536-01	METC49	127	0.0182 -		ISM02.4	PPB	5/8/2020 14:19	SMPL
L2536-02	METC49D	103	0.0127 -		ISM02.4	PPB	5/8/2020 14:22	SMPL
L2536-03	METC49S	4360	0.9858 -		ISM02.4	PPB	5/8/2020 14:27	SMPL
L2536-04	METC51	3	-0.0101 -		ISM02.4	PPB	5/8/2020 14:29	SMPL
L2536-05	METC53	11	-0.0083 -		ISM02.4	PPB	5/8/2020 14:31	SMPL
L2536-06	METC55	19	-0.0065 -		ISM02.4	PPB	5/8/2020 14:34	SMPL
CCV094	CCV094	21599	4.9261 -		ISM02.4	PPB	5/8/2020 14:36	SMPL
CCB094	CCB094	85	0.0086 -		ISM02.4	PPB	5/8/2020 14:38	SMPL