


LB109078

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3607e-004

C= -2.8342e-002

Rho= 0.9999587

Accept=Accepted

slope
+1 intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.00	0.000	-0.012	-0.012	70	0.000	70					11
0.05	0.050					223					-8
0.20	0.200	0.184	-0.016	899	0.0 %	899					1
2.50	2.500	2.516	0.016	10777	0.0 %	10777					1
5.00	5.000	5.059	0.059	21548	0.0 %	21548					1
7.50	7.500	7.449	-0.051	31672	0.0 %	31672					1
10.0	10.000	10.005	0.005	42501	0.0 %	42501					0

LB109078
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	70	-		0 ISM02.4	PPB	5/8/2020 15:54	Std
	0.2 S0.2	899	-		0.2 ISM02.4	PPB	5/8/2020 15:56	Std
	2.5 S2.5	10777	-		2.5 ISM02.4	PPB	5/8/2020 15:59	Std
	5 S5.0	21548	-		5 ISM02.4	PPB	5/8/2020 16:01	Std
	7.5 S7.5	31672	-		7.5 ISM02.4	PPB	5/8/2020 16:03	Std
	10 S10.0	42501	-		10 ISM02.4	PPB	5/8/2020 16:05	Std
ICV	ICV	16478	3.8617	-	ISM02.4	PPB	5/8/2020 16:08	SMPL
ICB	ICB	-68	-0.0444	-	ISM02.4	PPB	5/8/2020 16:10	SMPL
CCV095	CCV095	21342	5.0099	-	ISM02.4	PPB	5/8/2020 16:13	SMPL
CCB095	CCB095	-70	-0.0449	-	ISM02.4	PPB	5/8/2020 16:15	SMPL
PB128757BL	PBW757	13	-0.0253	-	ISM02.4	PPB	5/8/2020 16:17	SMPL
L2558-01	MHOKD0	16	-0.0246	-	ISM02.4	PPB	5/8/2020 16:19	SMPL
L2558-02	MHOKD2	14	-0.025	-	ISM02.4	PPB	5/8/2020 16:22	SMPL
L2558-03	MHOKD4	18	-0.0241	-	ISM02.4	PPB	5/8/2020 16:24	SMPL
L2558-04	MHOKD6	3	-0.0276	-	ISM02.4	PPB	5/8/2020 16:26	SMPL
L2558-05	MHOKD6D	-2	-0.0288	-	ISM02.4	PPB	5/8/2020 16:29	SMPL
L2558-06	MHOKD6S	4711	1.0838	-	ISM02.4	PPB	5/8/2020 16:33	SMPL
L2558-07	MHOKD8	-27	-0.0347	-	ISM02.4	PPB	5/8/2020 16:36	SMPL
L2558-08	MHOKEO	-29	-0.0352	-	ISM02.4	PPB	5/8/2020 16:38	SMPL
L2558-09	MHOKKE2	3	-0.0276	-	ISM02.4	PPB	5/8/2020 16:40	SMPL
L2558-10	MHOKKE4	-18	-0.0326	-	ISM02.4	PPB	5/8/2020 16:42	SMPL
L2558-11	MHOKKE6	-11	-0.0309	-	ISM02.4	PPB	5/8/2020 16:45	SMPL
L2558-12	MHOKKE8	-7	-0.03	-	ISM02.4	PPB	5/8/2020 16:47	SMPL
L2558-13	MHOKF1	-19	-0.0328	-	ISM02.4	PPB	5/8/2020 16:49	SMPL
L2558-14	MHOKF3	-20	-0.0331	-	ISM02.4	PPB	5/8/2020 16:52	SMPL
L2558-15	MHOKF5	-14	-0.0316	-	ISM02.4	PPB	5/8/2020 16:54	SMPL
L2558-16	MHOKF7	-27	-0.0347	-	ISM02.4	PPB	5/8/2020 16:56	SMPL
L2558-17	MHOKF9	-9	-0.0305	-	ISM02.4	PPB	5/8/2020 16:58	SMPL
L2558-18	MHOKG1	-9	-0.0305	-	ISM02.4	PPB	5/8/2020 17:01	SMPL
L2558-19	MHOKG3	-11	-0.0309	-	ISM02.4	PPB	5/8/2020 17:03	SMPL
CCV096	CCV096	21529	5.0541	-	ISM02.4	PPB	5/8/2020 17:05	SMPL
CCB096	CCB096	-71	-0.0451	-	ISM02.4	PPB	5/8/2020 17:07	SMPL
L2558-20	MHOKG5	-3	-0.029	-	ISM02.4	PPB	5/8/2020 17:10	SMPL
L2558-21	MHOKG7	17	-0.0243	-	ISM02.4	PPB	5/8/2020 17:12	SMPL
L2558-22	MHOKG9	8	-0.0265	-	ISM02.4	PPB	5/8/2020 17:14	SMPL
PB128758BL	PBW758	41	-0.0187	-	ISM02.4	PPB	5/8/2020 17:16	SMPL
L2560-01	MHOKD1	39	-0.0191	-	ISM02.4	PPB	5/8/2020 17:19	SMPL
L2560-02	MHOKD3	37	-0.0196	-	ISM02.4	PPB	5/8/2020 17:21	SMPL
L2560-03	MHOKD5	24	-0.0227	-	ISM02.4	PPB	5/8/2020 17:23	SMPL
L2560-04	MHOKD7	137	0.004	-	ISM02.4	PPB	5/8/2020 17:25	SMPL
L2560-05	MHOKD7D	100	-0.0047	-	ISM02.4	PPB	5/8/2020 17:28	SMPL
L2560-06	MHOKD7S	4625	1.0635	-	ISM02.4	PPB	5/8/2020 17:30	SMPL
L2560-07	MHOKD9	10	-0.026	-	ISM02.4	PPB	5/8/2020 17:32	SMPL
L2560-08	MHOKKE1	-6	-0.0298	-	ISM02.4	PPB	5/8/2020 17:35	SMPL
L2560-09	MHOKKE3	-17	-0.0324	-	ISM02.4	PPB	5/8/2020 17:37	SMPL
L2560-10	MHOKKE5	-10	-0.0307	-	ISM02.4	PPB	5/8/2020 17:39	SMPL

LB109078

INSTRUMENT ID : CV1

L2560-11	MH0KE7	6	-0.0269 -	ISM02.4	PPB	5/8/2020 17:41 SMPL
L2560-12	MH0KE9	17	-0.0243 -	ISM02.4	PPB	5/8/2020 17:44 SMPL
L2560-13	MH0KF0	-21	-0.0333 -	ISM02.4	PPB	5/8/2020 17:46 SMPL
L2560-14	MH0KF2	-16	-0.0321 -	ISM02.4	PPB	5/8/2020 17:48 SMPL
L2560-15	MH0KF4	-17	-0.0324 -	ISM02.4	PPB	5/8/2020 17:50 SMPL
L2560-16	MH0KF6	-5	-0.0295 -	ISM02.4	PPB	5/8/2020 17:53 SMPL
CCV097	CCV097	21106	4.9542 -	ISM02.4	PPB	5/8/2020 17:55 SMPL
CCB097	CCB097	-69	-0.0446 -	ISM02.4	PPB	5/8/2020 17:57 SMPL
L2560-17	MH0KF8	-28	-0.035 -	ISM02.4	PPB	5/8/2020 17:59 SMPL
L2560-18	MH0KG0	-2	-0.0288 -	ISM02.4	PPB	5/8/2020 18:02 SMPL
L2560-19	MH0KG2	0	-0.0283 -	ISM02.4	PPB	5/8/2020 18:04 SMPL
L2560-20	MH0KG4	17	-0.0243 -	ISM02.4	PPB	5/8/2020 18:06 SMPL
L2560-21	MH0KG6	-22	-0.0335 -	ISM02.4	PPB	5/8/2020 18:09 SMPL
L2560-22	MH0KG8	0	-0.0283 -	ISM02.4	PPB	5/8/2020 18:11 SMPL
CCV098	CCV098	20932	4.9131 -	ISM02.4	PPB	5/8/2020 18:13 SMPL
CCB098	CCB098	-65	-0.0437 -	ISM02.4	PPB	5/8/2020 18:15 SMPL