



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109080

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/08/2020	16:40
2	CAL2	1	Water	NA	NA	20.1	7.00	05/08/2020	16:41
3	CAL3	1	Water	NA	NA	20.1	10.03	05/08/2020	16:42
4	ICV	1	Water	NA	NA	20.1	7.00	05/08/2020	16:43
5	CCV1	1	Water	NA	NA	20.1	2.01	05/08/2020	16:45
6	L2502-06	1	Solid	20.03	20	20.2	7.99	05/08/2020	16:46
7	L2502-06DUP	1	Solid	20.04	20	20.3	8.00	05/08/2020	16:47
8	L2571-01	1	Solid	20.02	20	19.7	7.12	05/08/2020	16:49
9	L2571-02	1	Solid	20.03	20	18.8	5.76	05/08/2020	16:50
10	L2571-03	1	Solid	20.04	20	19.5	6.75	05/08/2020	17:00
11	L2571-04	1	Solid	20.02	20	19.8	6.56	05/08/2020	17:02
12	L2571-05	1	Solid	20.03	20	19.7	6.53	05/08/2020	17:05
13	CCV2	1	Water	NA	NA	20.2	12.02	05/08/2020	17:06