

LB109094


245.1

INSTRUMENT ID: CX2

Linear

 μ Abs.:

39489



A= 0.0000e+000

B= 2.5427e-004

C= 4.4311e-004

Rho= 0.9999752

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	4	0.000	4				
0.05	0.050	0.058	0.008	227	0.0 %	227				
0.2	0.200	0.201	0.001	790	0.0 %	790				
2.5	2.500	2.510	0.010	9871	0.0 %	9871				
5.0	5.000	4.990	-0.010	19623	0.0 %	19623				
7.5	7.500	7.447	-0.053	29286	0.0 %	29286				
10.0	10.000	10.041	0.041	39489	0.0 %	39489				

LB109094 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 May 2020 12:10:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	245.1	PPB	11 May 2020 12:14:01	S	Std
0.05 - 1	50.05	227	-	0.0500	245.1	PPB	11 May 2020 12:16:17	S	Std
0.2 - 1	50.2	790	-	0.2000	245.1	PPB	11 May 2020 12:18:35	S	Std
2.5 - 1	52.5	9871	-	2.5000	245.1	PPB	11 May 2020 12:20:52	S	Std
5.0 - 1	55	19623	-	5.0000	245.1	PPB	11 May 2020 12:23:08	S	Std
7.5 - 1	57.5	29286	-	7.5000	245.1	PPB	11 May 2020 12:25:24	S	Std
10.0 - 1	57.0	39489	-	10.0000	245.1	PPB	11 May 2020 12:43:23	S	Std
ICV39 - 1	ICV39	15798	4.0175	-245.1	PPB	PPB	11 May 2020 12:52:50	U	SMPL
ICB39 - 1	ICB39	-157	-0.0395	-245.1	PPB	PPB	11 May 2020 12:55:04	U	SMPL
CCV04 - 1	CCV04	19891	5.0582	-245.1	PPB	PPB	11 May 2020 12:57:20	U	SMPL
CCB04 - 1	CCB04	-147	-0.0369	-245.1	PPB	PPB	11 May 2020 12:59:35	U	SMPL
CRA - 1	CRA	169	0.0434	-245.1	PPB	PPB	11 May 2020 13:01:50	U	SMPL
HighStd - 1	HighStd	38910	9.8943	-245.1	PPB	PPB	11 May 2020 13:04:06	U	SMPL
ChkStd - 1	ChkStd	27117	6.8956	-245.1	PPB	PPB	11 May 2020 13:06:21	U	SMPL
PB128784BL - 1	PBW	-53	-0.0130	-245.1	PPB	PPB	11 May 2020 13:11:19	U	SMPL
PB128784BS - 1	LCSW	14404	3.6630	-245.1	PPB	PPB	11 May 2020 13:16:07	U	SMPL
L2113-01 - 1	COMP	88	0.0228	-245.1	PPB	PPB	11 May 2020 13:18:23	U	SMPL
L2113-01DUP - 1	COMPDUP	59	0.0154	-245.1	PPB	PPB	11 May 2020 13:20:39	U	SMPL
L2113-01MS - 1	COMPMS	13585	3.4548	-245.1	PPB	PPB	11 May 2020 13:22:56	U	SMPL
L2113-01MSD - 1	COMPMSD	14052	3.5735	-245.1	PPB	PPB	11 May 2020 13:25:13	U	SMPL
L2573-03 - 1	TI-604	-76	-0.0189	-245.1	PPB	PPB	11 May 2020 13:27:29	U	SMPL
CCV05 - 1	CCV05	20123	5.1172	-245.1	PPB	PPB	11 May 2020 13:29:44	U	SMPL
CCB05 - 1	CCB05	-137	-0.0344	-245.1	PPB	PPB	11 May 2020 13:32:00	U	SMPL
L2574-03 - 1	HP-640	26	0.0071	-245.1	PPB	PPB	11 May 2020 13:34:16	U	SMPL
L2113-01LX5 - 1		-6	-0.0011	-245.1	PPB	PPB	11 May 2020 13:36:31	U	SMPL
L2113-01A - 1		13800	3.5094	-245.1	PPB	PPB	11 May 2020 13:38:46	U	SMPL
CCV06 - 1	CCV06	19750	5.0224	-245.1	PPB	PPB	11 May 2020 13:41:02	U	SMPL
CCB06 - 1	CCB06	-149	-0.0374	-245.1	PPB	PPB	11 May 2020 13:43:18	U	SMPL