



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/12/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 05/11/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 05/12/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB109095

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2575-01	TP-28	18	1.11	9.61	8.44	86.2	
L2575-02	TP-28	19	1.14	9.97	9.12	90.4	
L2575-04	TP-28-EPH-2	20	1.17	9.94	8.62	84.9	
L2575-05	TP-29	21	1.12	9.59	8.39	85.8	
L2575-06	TP-29	22	1.16	9.98	8.83	87.0	
L2575-08	TP-29-EPH-2	23	1.18	9.77	8.48	85.0	
L2575-09	TP-27	24	1.13	9.69	9.15	93.7	
L2575-10	TP-27	25	1.19	9.7	8.61	87.2	
L2575-12	TP-27-EPH-2	26	1.12	9.62	8.45	86.2	
L2576-01	551-A2-DRUMS-COMP	27	1.00	2.00	2.00	100.0	SUB
L2580-01	400508030.0	1	1.13	9.88	5.32	47.9	sludge sample
L2581-01	400508031.0	2	1.16	9.83	1.66	5.8	sludge sample
L2582-01	400505704.0	3	1.15	9.94	1.4	2.8	sludge sample
L2583-01	400505767.0	4	1.19	9.9	4.14	33.9	sludge sample
L2584-01	400507706.0	5	1.14	9.7	1.82	7.9	sludge sample
L2585-01	400506707.1	6	1.13	9.74	2.61	17.2	sludge sample
L2586-01	400507717.0	7	1.17	9.72	2.67	17.5	sludge sample
L2587-01	400507771.0	8	1.19	9.97	3.15	22.3	sludge sample
L2588-01	400506791.0	9	1.1	9.73	3.12	23.4	sludge sample
L2593-01	FK-GF-02-057-A	10	1.18	9.73	9.05	92.0	
L2593-02	FK-GF-02-057-A	11	1.16	9.57	8.92	92.3	
L2593-03	FK-GF-02-057-B	12	1.18	9.67	8.99	92.0	
L2593-04	FK-GF-02-057-B	13	1.11	9.72	9.06	92.3	
L2594-01	FK-GF-02-056-A	14	1.13	9.58	8.88	91.7	
L2594-02	FK-GF-02-056-A	15	1.19	9.59	8.93	92.1	
L2594-03	FK-GF-02-056-B	16	1.15	9.83	9.14	92.1	
L2594-04	FK-GF-02-056-B	17	1.11	9.9	9.22	92.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$