


LB109096

INSTRUMENT ID: CY1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	28	0.000	28				
0.2	0.200	0.193	-0.007	788	0.0 %	788				
2.5	2.500	2.526	0.026	9857	0.0 %	9857				
5.0	5.000	4.921	-0.079	19171	0.0 %	19171				
7.5	7.500	7.625	0.125	29682	0.0 %	29682				
10.0	10.000	9.937	-0.063	38673	0.0 %	38673				

LB109096 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 May 2020 14:56:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.00007470A	PPB		11 May 2020 14:57:21	S	Std
0.2 - 1	50.2	788	-	0.20007470A	PPB		11 May 2020 14:59:29	S	Std
2.5 - 1	52.5	9857	-	2.50007470A	PPB		11 May 2020 15:01:37	S	Std
5.0 - 1	55	19171	-	5.00007470A	PPB		11 May 2020 15:03:44	S	Std
7.5 - 1	57.5	29682	-	7.50007470A	PPB		11 May 2020 15:05:51	S	Std
10.0 - 1	57.5	38673	-	10.00007470A	PPB		11 May 2020 15:07:58	S	Std
ICV40 - 1	ICV40	14567	3.7372	-7470A	PPB		11 May 2020 15:10:34	U	SMPL
ICB40 - 1	ICB40	-141	-0.0455	-7470A	PPB		11 May 2020 15:12:39	U	SMPL
CCV07 - 1	CCV07	19238	4.9385	-7470A	PPB		11 May 2020 15:14:45	U	SMPL
CCB07 - 1	CCB07	-148	-0.0473	-7470A	PPB		11 May 2020 15:16:50	U	SMPL
CRA - 1	CRA	747	0.1829	-7470A	PPB		11 May 2020 15:18:56	U	SMPL
HighStd - 1	HighStd	38674	9.9372	-7470A	PPB		11 May 2020 15:21:01	U	SMPL
ChkStd - 1	ChkStd	26878	6.9034	-7470A	PPB		11 May 2020 15:23:07	U	SMPL
PB128779BL - 1	PBW	-196	-0.0597	-7470A	PPB		11 May 2020 15:25:15	U	SMPL
PB128779BS - 1	LCSW	15094	3.8727	-7470A	PPB		11 May 2020 15:27:22	U	SMPL
L2502-06 - 1	PL-01-050820	-137	-0.0445	-7470A	PPB		11 May 2020 15:29:29	U	SMPL
L2502-06DUP - 1	PL-01-050820DUP	-30	-0.0170	-7470A	PPB		11 May 2020 15:31:37	U	SMPL
L2502-06MS - 1	PL-01-050820MS	14253	3.6564	-7470A	PPB		11 May 2020 15:33:44	U	SMPL
L2502-06MSD - 1	PL-01-050820MSD	15373	3.9445	-7470A	PPB		11 May 2020 15:35:52	U	SMPL
L2562-01 - 1	CHF-PAINT	-98	-0.0345	-7470A	PPB		11 May 2020 15:37:58	U	SMPL
CCV08 - 1	CCV08	19226	4.9354	-7470A	PPB		11 May 2020 15:40:04	U	SMPL
CCB08 - 1	CCB08	-147	-0.0471	-7470A	PPB		11 May 2020 15:42:10	U	SMPL
L2571-01 - 1	CHASSIE-WASH	226	0.0489	-7470A	PPB		11 May 2020 15:44:17	U	SMPL
L2571-02 - 1	OIL-WATER-SEPARATOR	-76	-0.0288	-7470A	PPB		11 May 2020 15:46:23	U	SMPL
L2571-03 - 1	BUS-LANE-1	84	0.0123	-7470A	PPB		11 May 2020 15:48:29	U	SMPL
L2571-04 - 1	BUS-LANE-3	28	-0.0021	-7470A	PPB		11 May 2020 15:50:35	U	SMPL
L2571-05 - 1	FUEL-LANE	0	-0.0093	-7470A	PPB		11 May 2020 15:52:41	U	SMPL
PB128780BL - 1	PBW	-28	-0.0165	-7470A	PPB		11 May 2020 15:54:49	U	SMPL
PB128780BS - 1	LCSW	15608	4.0049	-7470A	PPB		11 May 2020 15:56:56	U	SMPL
L2492-01 - 1	OIL-WATER-COMP	-71	-0.0275	-7470A	PPB		11 May 2020 15:59:04	U	SMPL
L2492-01DUP - 1	OIL-WATER-COMPDUP	-6	-0.0108	-7470A	PPB		11 May 2020 16:01:12	U	SMPL
L2492-01MS - 1	OIL-WATER-COMPMS	15429	3.9589	-7470A	PPB		11 May 2020 16:03:18	U	SMPL
CCV09 - 1	CCV09	19043	4.8884	-7470A	PPB		11 May 2020 16:05:24	U	SMPL
CCB09 - 1	CCB09	-132	-0.0432	-7470A	PPB		11 May 2020 16:07:31	U	SMPL
L2492-01MSD - 1	OIL-WATER-COMPMSD	15308	3.9278	-7470A	PPB		11 May 2020 16:09:37	U	SMPL
PB128779TB - 1	PB128760TB	-104	-0.0360	-7470A	PPB		11 May 2020 16:11:43	U	SMPL
PB128780TB - 1	PB128776TB	-29	-0.0167	-7470A	PPB		11 May 2020 16:13:49	U	SMPL
L2502-06LX5 - 1		-13	-0.0126	-7470A	PPB		11 May 2020 16:15:56	U	SMPL
L2502-06A - 1		15385	3.9476	-7470A	PPB		11 May 2020 16:18:02	U	SMPL
L2492-01LX5 - 1		-122	-0.0406	-7470A	PPB		11 May 2020 16:20:09	U	SMPL
L2492-01A - 1		15604	4.0039	-7470A	PPB		11 May 2020 16:22:17	U	SMPL
CCV10 - 1	CCV10	19205	4.9300	-7470A	PPB		11 May 2020 16:24:24	U	SMPL
CCB10 - 1	CCB10	-144	-0.0463	-7470A	PPB		11 May 2020 16:26:32	U	SMPL