



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109106

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/11/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	05/11/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	05/11/2020	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	05/11/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	05/11/2020	17:06
6	L2576-04	1	Water	NA	NA	14.8	7.31	05/11/2020	17:20
7	L2576-04DUP	1	Water	NA	NA	14.9	7.32	05/11/2020	17:21
8	CCV2	1	Water	NA	NA	20.2	12.02	05/12/2020	17:22

VB109106

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2576 WorkList ID : 138621 Department : Wet-Chemistry Date : 05-11-2020 16:34:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2020	Water	L2576-04 -C	Corrosivity	Cool 4 deg C	PSEG01	M13	FRAC-TANK-F05194	05/11/2020	9040C

Date/Time 05-11-2020 16:40
Received by: Jo SP
Relinquished by:

Date/Time 05-11-2020 17:40
Received by: SP
Relinquished by: Jn