

LB109112


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43976



A= 0.0000e+000

B= 2.2561e-004

C= -7.9871e-003

Rho= 0.9998429

Accept=Accepted

90% slope
7-12/12/2012

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	11	0.000	11					
0.05	0.050					223					
0.20	0.200	0.180	-0.020	832	0.0 %	832					-10
2.50	2.500	2.515	0.015	11185	0.0 %	11185					-1
5.00	5.000	4.969	-0.031	22058	0.0 %	22058					-1
7.50	7.500	7.628	0.128	33848	0.0 %	33848					2
10.0	10.000	9.913	-0.087	43976	0.0 %	43976					-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	11	-		0 ISM02.4	PPB	5/12/2020 13:32	Std
	0.2 S0.2	832	-		0.2 ISM02.4	PPB	5/12/2020 13:34	Std
	2.5 S2.5	11185	-		2.5 ISM02.4	PPB	5/12/2020 13:36	Std
	5 SS.0	22058	-		5 ISM02.4	PPB	5/12/2020 13:39	Std
	7.5 S7.5	33848	-		7.5 ISM02.4	PPB	5/12/2020 13:41	Std
	10 S10.0	43976	-		10 ISM02.4	PPB	5/12/2020 13:43	Std
ICV	ICV	16883	3.801 -		ISM02.4	PPB	5/12/2020 13:46	SMPL
ICB	ICB	-128	-0.0369 -		ISM02.4	PPB	5/12/2020 13:48	SMPL
CCV017	CCV017	22031	4.9624 -		ISM02.4	PPB	5/12/2020 13:51	SMPL
CCB017	CCB017	-131	-0.0375 -		ISM02.4	PPB	5/12/2020 13:53	SMPL
PB128818BL	PBW818	187	0.0342 -		ISM02.4	PPB	5/12/2020 13:55	SMPL
L2559-01	MH0KH1	121	0.0193 -		ISM02.4	PPB	5/12/2020 13:57	SMPL
L2559-02	MH0KH3	121	0.0193 -		ISM02.4	PPB	5/12/2020 14:00	SMPL
L2559-03	MH0KH5	115	0.018 -		ISM02.4	PPB	5/12/2020 14:02	SMPL
L2559-04	MH0KH7	114	0.0177 -		ISM02.4	PPB	5/12/2020 14:04	SMPL
L2559-05	MH0KH9	126	0.0204 -		ISM02.4	PPB	5/12/2020 14:06	SMPL
L2559-06	MH0KJ1	130	0.0213 -		ISM02.4	PPB	5/12/2020 14:09	SMPL
L2559-07	MH0KJ3	117	0.0184 -		ISM02.4	PPB	5/12/2020 14:11	SMPL
L2559-08	MH0KJ5	137	0.0229 -		ISM02.4	PPB	5/12/2020 14:13	SMPL
L2559-09	MH0KJ7	118	0.0186 -		ISM02.4	PPB	5/12/2020 14:15	SMPL
L2559-10	MH0KJ9	132	0.0218 -		ISM02.4	PPB	5/12/2020 14:18	SMPL
L2559-11	MH0KK1	134	0.0222 -		ISM02.4	PPB	5/12/2020 14:20	SMPL
L2559-12	MH0KK3	143	0.0243 -		ISM02.4	PPB	5/12/2020 14:22	SMPL
L2559-13	MH0KK5	128	0.0209 -		ISM02.4	PPB	5/12/2020 14:24	SMPL
L2559-14	MH0KK7	124	0.02 -		ISM02.4	PPB	5/12/2020 14:27	SMPL
L2559-15	MH0KK9	124	0.02 -		ISM02.4	PPB	5/12/2020 14:29	SMPL
L2559-16	MH0KL1	125	0.0202 -		ISM02.4	PPB	5/12/2020 14:31	SMPL
L2559-17	MH0KL3	116	0.0182 -		ISM02.4	PPB	5/12/2020 14:34	SMPL
L2559-18	MH0KL3D	112	0.0173 -		ISM02.4	PPB	5/12/2020 14:36	SMPL
L2559-19	MH0KL3S	4518	1.0113 -		ISM02.4	PPB	5/12/2020 14:38	SMPL
CCV018	CCV018	21764	4.9022 -		ISM02.4	PPB	5/12/2020 14:40	SMPL
CCB018	CCB018	-123	-0.0357 -		ISM02.4	PPB	5/12/2020 14:43	SMPL
PB128819BL	PBW819	186	0.034 -		ISM02.4	PPB	5/12/2020 14:45	SMPL
L2561-01	MH0KH0	-49	-0.019 -		ISM02.4	PPB	5/12/2020 14:47	SMPL
L2561-02	MH0KH2	-38	-0.0166 -		ISM02.4	PPB	5/12/2020 14:49	SMPL
L2561-03	MH0KH4	-32	-0.0152 -		ISM02.4	PPB	5/12/2020 14:52	SMPL
L2561-04	MH0KH6	-46	-0.0184 -		ISM02.4	PPB	5/12/2020 14:54	SMPL
L2561-05	MH0KH8	-53	-0.0199 -		ISM02.4	PPB	5/12/2020 14:56	SMPL
L2561-06	MH0KJ0	-39	-0.0168 -		ISM02.4	PPB	5/12/2020 14:58	SMPL
L2561-07	MH0KJ2	-40	-0.017 -		ISM02.4	PPB	5/12/2020 15:01	SMPL
L2561-08	MH0KJ4	-50	-0.0193 -		ISM02.4	PPB	5/12/2020 15:03	SMPL
L2561-09	MH0KJ6	-39	-0.0168 -		ISM02.4	PPB	5/12/2020 15:05	SMPL
L2561-10	MH0KJ8	-41	-0.0172 -		ISM02.4	PPB	5/12/2020 15:08	SMPL
L2561-11	MH0KK0	-25	-0.0136 -		ISM02.4	PPB	5/12/2020 15:10	SMPL
L2561-12	MH0KK2	-38	-0.0166 -		ISM02.4	PPB	5/12/2020 15:12	SMPL
L2561-13	MH0KK4	-32	-0.0152 -		ISM02.4	PPB	5/12/2020 15:14	SMPL

LB109112

INSTRUMENT ID : CV1

L2561-14	MHOKK6	-45	-0.0181 -	ISM02.4	PPB	5/12/2020 15:17 SMPL
L2561-15	MHOKK8	-52	-0.0197 -	ISM02.4	PPB	5/12/2020 15:19 SMPL
L2561-16	MHOKL0	-46	-0.0184 -	ISM02.4	PPB	5/12/2020 15:21 SMPL
L2561-17	MHOKL2	-47	-0.0186 -	ISM02.4	PPB	5/12/2020 15:23 SMPL
L2561-18	MHOKL4	-47	-0.0186 -	ISM02.4	PPB	5/12/2020 15:26 SMPL
L2561-19	MHOKL4D	-40	-0.017 -	ISM02.4	PPB	5/12/2020 15:28 SMPL
CCV019	CCV019	21544	4.8525 -	ISM02.4	PPB	5/12/2020 15:30 SMPL
CCB019	CCB019	-122	-0.0355 -	ISM02.4	PPB	5/12/2020 15:32 SMPL
L2561-20	MHOKL4S	4476	1.0018 -	ISM02.4	PPB	5/12/2020 15:35 SMPL
PB128820BL	PBW820	-46	-0.0184 -	ISM02.4	PPB	5/12/2020 15:37 SMPL
L2577-01	MBFSF4	136	0.0227 -	ISM02.4	PPB	5/12/2020 15:39 SMPL
L2577-02	MBFSF4D	22	-0.003 -	ISM02.4	PPB	5/12/2020 15:42 SMPL
L2577-03	MBFSF4S	4466	0.9996 -	ISM02.4	PPB	5/12/2020 15:46 SMPL
L2577-04	MBFSF5	-7	-0.0096 -	ISM02.4	PPB	5/12/2020 15:49 SMPL
CCV020	CCV020	21743	4.8974 -	ISM02.4	PPB	5/12/2020 15:51 SMPL
CCB020	CCB020	-101	-0.0308 -	ISM02.4	PPB	5/12/2020 15:53 SMPL