



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109114

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	05/12/2020	14:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/12/2020	14:01
3	CAL3	1	Water	NA	NA	20.2	10.03	05/12/2020	14:02
4	ICV	1	Water	NA	NA	20.1	7.02	05/12/2020	14:05
5	CCV1	1	Water	NA	NA	20.2	2.02	05/12/2020	14:06
6	L2600-01	1	Water	NA	NA	15.8	11.58	05/12/2020	14:10
7	L2600-01DUP	1	Water	NA	NA	15.9	11.59	05/12/2020	14:11
8	L2600-02	1	Water	NA	NA	17.1	11.56	05/12/2020	14:13
9	L2600-03	1	Water	NA	NA	16.9	11.53	05/12/2020	14:15
10	CCV2	1	Water	NA	NA	20.2	12.02	05/12/2020	14:16

VB109114

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2600 WorkList ID : 138665 Department : Wet-Chemistry Date : 05-12-2020 13:47:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/12/2020	Water	L2600-01	pH ✓	Cool 4 deg C	PSEG01	M23	VB25984	05/12/2020	9040C
05/12/2020	Water	L2600-02	pH ✓	Cool 4 deg C	PSEG01	M23	VB16122	05/12/2020	9040C
05/12/2020	Water	L2600-03	pH ✓	Cool 4 deg C	PSEG01	M23	VB15242	05/12/2020	9040C

Date/Time 05.12.2020 13:40
Received by: JP
Relinquished by: KS

Date/Time 05.12.2020 15:15
Received by: KS
Relinquished by: JP