

LB109137


INSTRUMENT ID: CV4

7470A

Linear

 μ Abs.:

38136



A= 0.0000e+000

B= 2.5916e-004

C= 8.7000e-004

Rho= 0.9996033

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	0	0.000	0				
0.2	0.200	0.202	0.002	777	0.0 %	777				
2.5	2.500	2.496	-0.004	9628	0.0 %	9628				
5.0	5.000	4.909	-0.091	18938	0.0 %	18938				
7.5	7.500	7.708	0.208	29737	0.0 %	29737				
10.0	10.000	9.884	-0.116	38136	0.0 %	38136				

LB109137 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 May 2020 14:39:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7470A	PPB	13 May 2020 14:40:24	S	Std
0.2 - 1	50.2	777	-	0.2000	7470A	PPB	13 May 2020 14:42:32	S	Std
2.5 - 1	52.5	9628	-	2.5000	7470A	PPB	13 May 2020 14:44:40	S	Std
5.0 - 1	55	18938	-	5.0000	7470A	PPB	13 May 2020 14:46:47	S	Std
7.5 - 1	57.5	29737	-	7.5000	7470A	PPB	13 May 2020 14:48:54	S	Std
10.0 - 1	510	38136	-	10.0000	7470A	PPB	13 May 2020 14:53:26	S	Std
ICV45 - 1	ICV45	14398	3.7323	-7470A	PPB	PPB	13 May 2020 14:56:13	U	SMPL
ICB45 - 1	ICB45	-137	-0.0346	-7470A	PPB	PPB	13 May 2020 14:58:19	U	SMPL
CCV29 - 1	CCV29	18584	4.8172	-7470A	PPB	PPB	13 May 2020 15:00:24	U	SMPL
CCB29 - 1	CCB29	-139	-0.0352	-7470A	PPB	PPB	13 May 2020 15:02:30	U	SMPL
CRA - 1	CRA	776	0.2020	-7470A	PPB	PPB	13 May 2020 15:04:36	U	SMPL
HighStd - 1	HighStd	38557	9.9934	-7470A	PPB	PPB	13 May 2020 15:06:42	U	SMPL
ChkStd - 1	ChkStd	26147	6.7772	-7470A	PPB	PPB	13 May 2020 15:08:48	U	SMPL
PB128859BL - 1	PBW	-196	-0.0499	-7470A	PPB	PPB	13 May 2020 15:10:56	U	SMPL
PB128859BS - 1	LCSW	15645	4.0555	-7470A	PPB	PPB	13 May 2020 15:13:04	U	SMPL
L2500-05 - 1	OK-02-051220	-135	-0.0341	-7470A	PPB	PPB	13 May 2020 15:15:11	U	SMPL
L2500-05DUP - 1	OK-02-051220DUP	-53	-0.0129	-7470A	PPB	PPB	13 May 2020 15:17:18	U	SMPL
L2500-05MS - 1	OK-02-051220MS	15644	4.0552	-7470A	PPB	PPB	13 May 2020 15:24:31	U	SMPL
L2500-05MSD - 1	OK-02-051220MSD	15606	4.0454	-7470A	PPB	PPB	13 May 2020 15:26:38	U	SMPL
L2505-05 - 1	SU-02-051220	-122	-0.0307	-7470A	PPB	PPB	13 May 2020 15:28:45	U	SMPL
CCV30 - 1	CCV30	18557	4.8102	-7470A	PPB	PPB	13 May 2020 15:30:51	U	SMPL
CCB30 - 1	CCB30	-149	-0.0377	-7470A	PPB	PPB	13 May 2020 15:32:57	U	SMPL
L2604-02 - 1	305	-38	-0.0090	-7470A	PPB	PPB	13 May 2020 15:35:03	U	SMPL
L2605-02 - 1	239992	-26	-0.0059	-7470A	PPB	PPB	13 May 2020 15:37:09	U	SMPL
L2605-05 - 1	239997	-20	-0.0043	-7470A	PPB	PPB	13 May 2020 15:39:16	U	SMPL
L2605-08 - 1	239996	-23	-0.0051	-7470A	PPB	PPB	13 May 2020 15:41:22	U	SMPL
PB128860BL - 1	PBW	-11	-0.0020	-7470A	PPB	PPB	13 May 2020 15:43:29	U	SMPL
PB128860BS - 1	LCSW	14874	3.8557	-7470A	PPB	PPB	13 May 2020 15:45:36	U	SMPL
L2596-01 - 1	R-1U	175	0.0462	-7470A	PPB	PPB	13 May 2020 15:57:16	U	SMPL
L2596-01DUP - 1	R-1UDUP	109	0.0291	-7470A	PPB	PPB	13 May 2020 15:59:23	U	SMPL
L2596-01MS - 1	R-1UMS	14636	3.7940	-7470A	PPB	PPB	13 May 2020 16:01:31	U	SMPL
L2596-01MSD - 1	R-1UMSD	14337	3.7165	-7470A	PPB	PPB	13 May 2020 16:03:37	U	SMPL
CCV31 - 1	CCV31	18813	4.8765	-7470A	PPB	PPB	13 May 2020 16:05:43	U	SMPL
CCB31 - 1	CCB31	-40	-0.0095	-7470A	PPB	PPB	13 May 2020 16:07:50	U	SMPL
L2596-02 - 1	E-1U	22	0.0066	-7470A	PPB	PPB	13 May 2020 16:09:56	U	SMPL
L2596-03 - 1	D-1U	307	0.0804	-7470A	PPB	PPB	13 May 2020 16:12:02	U	SMPL
L2596-04 - 1	R-1F	94	0.0252	-7470A	PPB	PPB	13 May 2020 16:14:09	U	SMPL
L2596-05 - 1	E-1F	5	0.0022	-7470A	PPB	PPB	13 May 2020 16:16:15	U	SMPL
L2596-06 - 1	D-1F	125	0.0333	-7470A	PPB	PPB	13 May 2020 16:18:21	U	SMPL
PB128859TB - 1	PB128837TB	50	0.0138	-7470A	PPB	PPB	13 May 2020 16:20:28	U	SMPL
L2500-05LX5 - 1		73	0.0198	-7470A	PPB	PPB	13 May 2020 16:22:35	U	SMPL
L2500-05A - 1		16336	4.2346	-7470A	PPB	PPB	13 May 2020 16:24:43	U	SMPL
L2596-01LX5 - 1		89	0.0239	-7470A	PPB	PPB	13 May 2020 16:29:33	U	SMPL
L2596-01A - 1		15663	4.0601	-7470A	PPB	PPB	13 May 2020 16:31:39	U	SMPL
CCV32 - 1	CCV32	18815	4.8770	-7470A	PPB	PPB	13 May 2020 16:33:46	U	SMPL
CCB32 - 1	CCB32	-85	-0.0212	-7470A	PPB	PPB	13 May 2020 16:35:52	U	SMPL