


LB109142

245.1 INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6765e-004

C= -1.6374e-002

Rho= 0.9999214

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	12	0.000	12				
0.05	0.050	0.040	-0.010	210	0.0 %	210				
0.2	0.200	0.187	-0.013	758	0.0 %	758				
2.5	2.500	2.545	0.045	9569	0.0 %	9569				
5.0	5.000	5.056	0.056	18950	0.0 %	18950				
7.5	7.500	7.407	-0.093	27734	0.0 %	27734				
10.0	10.000	10.030	0.030	37535	0.0 %	37535				

LB109142 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 May 2020 17:02:04

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	245.1	PPB	13 May 2020 17:04:27	S	Std
0.05 - 1	50.05	210	-	0.0500	245.1	PPB	13 May 2020 17:06:44	S	Std
0.2 - 1	50.2	758	-	0.2000	245.1	PPB	13 May 2020 17:09:01	S	Std
2.5 - 1	52.5	9569	-	2.5000	245.1	PPB	13 May 2020 17:11:17	S	Std
5.0 - 1	55	18950	-	5.0000	245.1	PPB	13 May 2020 17:13:33	S	Std
7.5 - 1	57.5	27734	-	7.5000	245.1	PPB	13 May 2020 17:15:49	S	Std
10.0 - 1	500	37535	-	10.0000	245.1	PPB	13 May 2020 17:20:45	S	Std
ICV46 - 1	ICV46	14351	3.8247	-245.1		PPB	13 May 2020 17:23:30	U	SMPL
ICB46 - 1	ICB46	-34	-0.0255	-245.1		PPB	13 May 2020 17:28:26	U	SMPL
CCV33 - 1	CCV33	18881	5.0371	-245.1		PPB	13 May 2020 17:30:40	U	SMPL
CCB33 - 1	CCB33	-81	-0.0381	-245.1		PPB	13 May 2020 17:32:56	U	SMPL
CRA - 1	CRA	230	0.0452	-245.1		PPB	13 May 2020 17:37:44	U	SMPL
HighStd - 1	HighStd	37866	10.1184	-245.1		PPB	13 May 2020 17:39:59	U	SMPL
ChkStd - 1	ChkStd	25665	6.8529	-245.1		PPB	13 May 2020 17:42:14	U	SMPL
PB128861BL - 1	PBW	-52	-0.0303	-245.1		PPB	13 May 2020 17:47:05	U	SMPL
PB128861BS - 1	LCSW	15421	4.1111	-245.1		PPB	13 May 2020 17:49:20	U	SMPL
L2591-01 - 1	3-11-27TH-AVE	204089	54.6080	-245.1		PPB	13 May 2020 17:51:36	U	SMPL
L2591-01DUP - 1	3-11-27TH-AVEDUP	205517	54.9902	-245.1		PPB	13 May 2020 17:53:53	U	SMPL
L2591-01MS - 1	3-11-27TH-AVEMS	209935	56.1727	-245.1		PPB	13 May 2020 17:56:19	U	SMPL
L2591-01MSD - 1	3-11-27TH-AVEMSD	198526	53.1190	-245.1		PPB	13 May 2020 17:58:41	U	SMPL
CCV34 - 1	CCV34	18388	4.9052	-245.1		PPB	13 May 2020 18:01:05	U	SMPL
CCB34 - 1	CCB34	-153	-0.0573	-245.1		PPB	13 May 2020 18:03:27	U	SMPL
L2591-01LX5 - 1		37137	9.9233	-245.1		PPB	13 May 2020 18:05:42	U	SMPL
L2591-01A - 1		207596	55.5466	-245.1		PPB	13 May 2020 18:07:57	U	SMPL
L2591-01DLX10 - 1		18114	4.8318	-245.1		PPB	13 May 2020 18:10:14	U	SMPL
L2591-01DUPDLX10 - 1		19043	5.0805	-245.1		PPB	13 May 2020 18:12:34	U	SMPL
L2591-01MSDLX10 - 1		19036	5.0786	-245.1		PPB	13 May 2020 18:14:51	U	SMPL
L2591-01MSDDLX10 - 1		19281	5.1442	-245.1		PPB	13 May 2020 18:17:08	U	SMPL
L2591-01LDLX50 - 1		3584	0.9429	-245.1		PPB	13 May 2020 18:19:25	U	SMPL
L2591-01ADLX10 - 1		19309	5.1517	-245.1		PPB	13 May 2020 18:21:42	U	SMPL
CCV35 - 1	CCV35	18893	5.0403	-245.1		PPB	13 May 2020 18:23:59	U	SMPL
CCB35 - 1	CCB35	-49	-0.0295	-245.1		PPB	13 May 2020 18:26:17	U	SMPL