



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109144

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/13/2020	17:15
2	CAL2	1	Water	NA	NA	20.2	7.02	05/13/2020	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	05/13/2020	17:18
4	ICV	1	Water	NA	NA	20.1	7.00	05/13/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.01	05/13/2020	17:21
6	L2501-07	1	Solid	20.03	20	20.8	9.07	05/13/2020	17:22
7	L2501-07DUP	1	Solid	20.04	20	20.8	9.08	05/13/2020	17:23
8	L2635-02	1	Solid	20.02	20	20.8	9.35	05/13/2020	17:25
9	L2638-06	1	Solid	20.03	20	18.6	9.10	05/13/2020	17:30
10	L2640-01	1	Solid	20.02	20	20.9	5.56	05/13/2020	17:32
11	L2641-02	1	Solid	20.01	20	19.6	7.71	05/13/2020	17:33
12	L2641-03	1	Solid	20.03	20	19.4	6.27	05/13/2020	17:35
13	L2641-04	1	Solid	20.04	20	19.2	6.17	05/13/2020	17:40
14	CCV2	1	Water	NA	NA	20.2	12.02	05/13/2020	17:36

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2641

WorkList ID : 138704

Department : Wet-Chemistry

Date : 05-13-2020 15:04:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/13/2020	Solid	L2501-07	Corrosivity	Cool 4 deg C	PSEG05	M42	EO-01-051320	05/13/2020	9045D
05/13/2020	Solid	L2635-02	Corrosivity	Cool 4 deg C	PSEG01	M33	ROLL-OFF-P72-11933	05/13/2020	9045D
05/13/2020	Solid	L2638-06	Corrosivity	Cool 4 deg C	DIST01	M31	SUP-WZ2-C-1	05/13/2020	9045D
05/13/2020	Solid	L2640-01	Corrosivity	Cool 4 deg C	ARAM01	M41	WASTE-WATER-CAKE	05/13/2020	9045D
05/08/2020	Solid	L2641-02	Corrosivity	Cool 4 deg C	CLEA11		CHASSIS-WASH-2	05/08/2020	9045D
05/08/2020	Solid	L2641-03	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH	05/08/2020	9045D
05/08/2020	Solid	L2641-04	Corrosivity	Cool 4 deg C	CLEA11		OWS-BUS-WASH	05/08/2020	9045D

Date/Time 05-13-2020 15:10

Received by: JS

Relinquished by: JS

Date/Time 05-13-2020 18:20

Received by: KS

Relinquished by: JS