


LB 109148

ISM02.4

INSTRUMENT ID: CV1



slope
7.151e-004

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.010	-0.010	-54	0.000	-54					
0.05	0.050					223					
0.20	0.200	0.181	-0.019	781	0.0 %	781					
2.50	2.500	2.500	-0.000	10902	0.0 %	10902					
5.00	5.000	5.005	0.005	21834	0.0 %	21834					
7.50	7.500	7.617	0.117	33238	0.0 %	33238					
10.0	10.000	9.908	-0.092	43234	0.0 %	43234					

%D
= -10
0
0
2
-1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-54	-		0 ISM02.4	PPB	5/14/2020 11:53	Std
	0.2 S0.2	781	-		0.2 ISM02.4	PPB	5/14/2020 11:55	Std
	2.5 S2.5	10902	-		2.5 ISM02.4	PPB	5/14/2020 11:57	Std
	5 S5.0	21834	-		5 ISM02.4	PPB	5/14/2020 12:02	Std
	7.5 S7.5	33238	-		7.5 ISM02.4	PPB	5/14/2020 12:04	Std
	10 S10.0	43234	-		10 ISM02.4	PPB	5/14/2020 12:07	Std
ICV	ICV	18191	4.1699 -		ISM02.4	PPB	5/14/2020 12:12	SMPL
ICB	ICB	-140	-0.0302 -		ISM02.4	PPB	5/14/2020 12:14	SMPL
CCV036	CCV036	22011	5.0451 -		ISM02.4	PPB	5/14/2020 12:16	SMPL
CCB036	CCB036	-160	-0.0347 -		ISM02.4	PPB	5/14/2020 12:19	SMPL
PB128880BL	PBW880	30	0.0088 -		ISM02.4	PPB	5/14/2020 12:21	SMPL
L2632-01	MBF8M5	-27	-0.0043 -		ISM02.4	PPB	5/14/2020 12:23	SMPL
L2632-02	MBF8M6	-35	-0.0061 -		ISM02.4	PPB	5/14/2020 12:25	SMPL
L2632-03	MBF8M7	-27	-0.0043 -		ISM02.4	PPB	5/14/2020 12:28	SMPL
L2632-04	MBF8M8	-48	-0.0091 -		ISM02.4	PPB	5/14/2020 12:30	SMPL
L2632-05	MBF8M9	-47	-0.0089 -		ISM02.4	PPB	5/14/2020 12:32	SMPL
L2632-06	MBF8N1	93	0.0232 -		ISM02.4	PPB	5/14/2020 12:34	SMPL
L2632-07	MBF8N2	-49	-0.0093 -		ISM02.4	PPB	5/14/2020 12:37	SMPL
L2632-08	MBF8N3	-52	-0.01 -		ISM02.4	PPB	5/14/2020 12:39	SMPL
L2632-09	MBF8N4	296	0.0697 -		ISM02.4	PPB	5/14/2020 12:41	SMPL
L2632-10	MBF8N5	145	0.0351 -		ISM02.4	PPB	5/14/2020 12:43	SMPL
L2632-11	MBF8N6	-3	0.0012 -		ISM02.4	PPB	5/14/2020 12:46	SMPL
L2632-12	MBF8N7	-36	-0.0063 -		ISM02.4	PPB	5/14/2020 12:48	SMPL
CCV037	CCV037	22010	5.0449 -		ISM02.4	PPB	5/14/2020 12:55	SMPL
CCB037	CCB037	-113	-0.024 -		ISM02.4	PPB	5/14/2020 12:57	SMPL