


LB109149

7470A

INSTRUMENT ID: CV1

 μ Abs.:

37179



Linear

A= 0.0000e+000

B= 2.6906e-004

C= -2.1898e-002

Rho= 0.9997895

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	31	0.000	31				
0.2	0.200	0.176	-0.024	735	0.0 %	735				
2.5	2.500	2.472	-0.028	9269	0.0 %	9269				
5.0	5.000	5.162	0.162	19268	0.0 %	19268				
7.5	7.500	7.422	-0.078	27665	0.0 %	27665				
10.0	10.000	9.982	-0.018	37179	0.0 %	37179				

LB109148 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 May 2020 14:00:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7470A	PPB	14 May 2020 14:02:30	S	Std
0.2 - 1	50.2	735	-	0.2000	7470A	PPB	14 May 2020 14:04:37	S	Std
2.5 - 1	52.5	9269	-	2.5000	7470A	PPB	14 May 2020 14:06:45	S	Std
5.0 - 1	55	19268	-	5.0000	7470A	PPB	14 May 2020 14:08:53	S	Std
7.5 - 1	57.5	27665	-	7.5000	7470A	PPB	14 May 2020 14:11:00	S	Std
10.0 - 1	57.5	37179	-	10.0000	7470A	PPB	14 May 2020 14:15:38	S	Std
ICV47 - 1	ICV47	14107	3.7738	-	7470A	PPB	14 May 2020 14:18:08	U	SMPL
ICB47 - 1	ICB47	-103	-0.0496	-	7470A	PPB	14 May 2020 14:20:13	U	SMPL
CCV38 - 1	CCV38	19131	5.1255	-	7470A	PPB	14 May 2020 14:22:19	U	SMPL
CCB38 - 1	CCB38	-120	-0.0542	-	7470A	PPB	14 May 2020 14:24:24	U	SMPL
CRA - 1	CRA	686	0.1627	-	7470A	PPB	14 May 2020 14:26:30	U	SMPL
HighStd - 1	HighStd	37884	10.1712	-	7470A	PPB	14 May 2020 14:28:36	U	SMPL
ChkStd - 1	ChkStd	25881	6.9417	-	7470A	PPB	14 May 2020 14:30:41	U	SMPL
PB128890BL - 1	PBW	-181	-0.0706	-	7470A	PPB	14 May 2020 14:32:49	U	SMPL
PB128890BS - 1	LCSW	15492	4.1464	-	7470A	PPB	14 May 2020 14:34:57	U	SMPL
L2501-07 - 1	EO-01-051320	-109	-0.0512	-	7470A	PPB	14 May 2020 14:37:04	U	SMPL
L2501-07DUP - 1	EO-01-051320DUP	-22	-0.0278	-	7470A	PPB	14 May 2020 14:39:11	U	SMPL
L2501-07MS - 1	EO-01-051320MS	14332	3.8343	-	7470A	PPB	14 May 2020 14:41:19	U	SMPL
L2501-07MSD - 1	EO-01-051320MSD	14195	3.7974	-	7470A	PPB	14 May 2020 14:43:26	U	SMPL
L2635-02 - 1	ROLL-OFF-P72-11933	-110	-0.0515	-	7470A	PPB	14 May 2020 14:45:32	U	SMPL
CCV39 - 1	CCV39	19192	5.1419	-	7470A	PPB	14 May 2020 14:47:39	U	SMPL
CCB39 - 1	CCB39	-135	-0.0582	-	7470A	PPB	14 May 2020 14:49:45	U	SMPL
L2638-06 - 1	SUP-WZ2-C-1	-25	-0.0286	-	7470A	PPB	14 May 2020 14:51:51	U	SMPL
L2640-01 - 1	WASTE-WATER-CAKE	8	-0.0197	-	7470A	PPB	14 May 2020 14:53:57	U	SMPL
L2641-02 - 1	CHASSIS-WASH-2	97	0.0042	-	7470A	PPB	14 May 2020 14:56:03	U	SMPL
L2641-03 - 1	BUS-WASH	12	-0.0187	-	7470A	PPB	14 May 2020 14:58:10	U	SMPL
L2641-04 - 1	OWS-BUS-WASH	33	-0.0130	-	7470A	PPB	14 May 2020 15:00:16	U	SMPL
PB128890TB - 1	PB128865TB	-17	-0.0265	-	7470A	PPB	14 May 2020 15:02:23	U	SMPL
L2501-07LX5 - 1		-4	-0.0230	-	7470A	PPB	14 May 2020 15:04:30	U	SMPL
L2501-07A - 1		15132	4.0495	-	7470A	PPB	14 May 2020 15:06:38	U	SMPL
CCV40 - 1	CCV40	18598	4.9821	-	7470A	PPB	14 May 2020 15:08:46	U	SMPL
CCB40 - 1	CCB40	-119	-0.0539	-	7470A	PPB	14 May 2020 15:10:52	U	SMPL