



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109160

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/14/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	05/14/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	05/14/2020	17:18
4	ICV	1	Water	NA	NA	20.2	7.00	05/14/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.01	05/14/2020	17:22
6	L2648-02	1	Water	NA	NA	18.1	6.55	05/14/2020	17:23
7	L2648-02DUP	1	Water	NA	NA	18.2	6.56	05/14/2020	17:25
8	CCV2	1	Water	NA	NA	20.1	12.02	05/14/2020	17:26

18109160

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSION L2644 WorkList ID : 138739 Department : Wet-Chemistry Date : 05-14-2020 14:36:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/13/2020	Water	L2648-02	Corrosivity	Cool 4 deg C	PSEG01	M43	TP-B	05/13/2020	9040C

Date/Time 05-14-2020 16:00
Received by: JD
Relinquished by: JD

Date/Time 05-14-2020 17:20
Received by: JD
Relinquished by: JD