



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109161

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/14/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	05/14/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	05/14/2020	17:19
4	ICV	1	Water	NA	NA	20.2	7.00	05/14/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.01	05/14/2020	17:22
6	L2644-01	1	Water	NA	NA	16.9	7.49	05/14/2020	17:25
7	L2644-02	1	Water	NA	NA	16.6	8.08	05/14/2020	17:30
8	L2644-03	1	Water	NA	NA	16.8	9.37	05/14/2020	17:32
9	L2644-04	1	Water	NA	NA	16.9	9.11	05/14/2020	17:33
10	L2644-05	1	Water	NA	NA	17.2	8.88	05/14/2020	17:35
11	L2648-01	1	Water	NA	NA	18.1	6.55	05/14/2020	17:23
12	L2648-01DUP	1	Water	NA	NA	18.2	6.56	05/14/2020	17:25
13	L2663-05	1	Water	NA	NA	21.0	6.92	05/14/2020	17:35
14	L2663-06	1	Water	NA	NA	19.8	6.53	05/14/2020	17:36
15	L2663-07	1	Water	NA	NA	20.3	6.71	05/14/2020	17:39
16	CCV2	1	Water	NA	NA	20.1	12.02	05/14/2020	17:40
17	L2663-08	1	Water	NA	NA	20.9	7.12	05/14/2020	17:42
18	L2663-12	1	Water	NA	NA	21.2	7.78	05/14/2020	17:46
19	CCV3	1	Water	NA	NA	21.3	2.01	05/14/2020	17:48

VB 109161

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2644

WorkList ID : 138738

Department : Wet-Chemistry

Date : 05-14-2020 14:36:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/14/2020	Water	L2644-01	pH	Cool 4 deg C	PSEG01	M42	1697	05/14/2020	9040C
05/14/2020	Water	L2644-02	pH	Cool 4 deg C	PSEG01	M42	1698	05/14/2020	9040C
05/14/2020	Water	L2644-03	pH	Cool 4 deg C	PSEG01	M42	1699	05/14/2020	9040C
05/14/2020	Water	L2644-04	pH	Cool 4 deg C	PSEG01	M42	1700	05/14/2020	9040C
05/14/2020	Water	L2644-05	pH	Cool 4 deg C	PSEG01	M42	1701	05/14/2020	9040C
05/13/2020	Water	L2648-01	pH	Cool 4 deg C	PSEG01	M43	TP-B	05/13/2020	9040C
05/14/2020	Water	L2663-05	pH	Cool 4 deg C	PSEG01	M51	1378	05/14/2020	9040C
05/14/2020	Water	L2663-06	pH	Cool 4 deg C	PSEG01	M51	1381	05/14/2020	9040C
05/14/2020	Water	L2663-07	pH	Cool 4 deg C	PSEG01	M51	1369	05/14/2020	9040C
05/14/2020	Water	L2663-08	pH	Cool 4 deg C	PSEG01	M51	1366	05/14/2020	9040C
05/14/2020	Water	L2663-12	pH	Cool 4 deg C	PSEG01	M51	1359	05/14/2020	9040C

Date/Time 05-14-2020 16:00
 Received by: Jn
 Relinquished by: Jn

Date/Time 05-14-2020 18:20
 Received by: Jn
 Relinquished by: Jn