



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109162

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/14/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	05/14/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	05/14/2020	17:18
4	ICV	1	Water	NA	NA	20.2	7.00	05/14/2020	17:20
5	CCV1	1	Water	NA	NA	20.3	2.02	05/14/2020	17:50
6	L2663-09	1	Solid	20.03	20	21.2	5.31	05/14/2020	18:00
7	L2663-09DUP	1	Solid	20.04	20	21.3	5.32	05/14/2020	18:01
8	CCV2	1	Water	NA	NA	20.2	12.02	05/14/2020	18:03

20109162

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2644 S WorkList ID : 138747 Department : Wet-Chemistry Date : 05-14-2020 16:54:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/14/2020	Solid	L2663-09	pH	Cool 4 deg C	PSEG01	M51	1364	05/14/2020	9040C

Date/Time 05-14-2020 16:10
Received by: JA
Relinquished by: JA

Date/Time 05-14-2020 18:20
Received by: JA
Relinquished by: JA