



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109180

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/15/2020	11:20
2	L2648-02		1	8	100.00	>212.0	>212.0	05/15/2020	11:50
3	L2648-02DUP		1	8	100.00	>212.0	>212.0	05/15/2020	12:15
4	L2663-12		1	9	100.00	>212.0	>212.0	05/15/2020	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp51520 WorkList ID : 138782 Department : Wet-Chemistry Date : 05-15-2020 10:40:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/26/2020	Water	L2648-02	Flash Point	Cool 4 deg C	PSEG01	M43	TP-B	05/13/2020	1010A
05/21/2020	Water	L2663-12	Flash Point	Cool 4 deg C	PSEG01	M51	1359	05/14/2020	1010A

Date/Time 05/15/20 10:45
 Received by: JS
 Relinquished by: JS

Date/Time 05/15/20 12:55
 Received by: JS
 Relinquished by: JS