


LB109184

7470A

INSTRUMENT ID: CV1

 μ Abs.:

39193



Linear

A= 0.0000e+000

B= 2.5777e-004

C= -1.1737e-001

Rho= 0.9999839

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	529	0.000	529				
0.2	0.200	0.192	-0.008	1201	0.0 %	1201				
2.5	2.500	2.486	-0.014	10101	0.0 %	10101				
5.0	5.000	4.980	-0.020	19773	0.0 %	19773				
7.5	7.500	7.537	0.037	29696	0.0 %	29696				
10.0	10.000	9.985	-0.015	39193	0.0 %	39193				

LB109184 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 May 2020 13:40:15					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	529	-	0.00007470A	PPB		15 May 2020 13:53:19	S	Std
0.2 - 1	50-2	1201	-	0.20007470A	PPB		15 May 2020 13:55:27	S	Std
2.5 - 1	52.5	10101	-	2.50007470A	PPB		15 May 2020 13:57:35	S	Std
5.0 - 1	55	19773	-	5.00007470A	PPB		15 May 2020 13:59:43	S	Std
7.5 - 1	57.5	29696	-	7.50007470A	PPB		15 May 2020 14:01:49	S	Std
10.0 - 1	510	39193	-	10.00007470A	PPB		15 May 2020 14:06:31	S	Std
ICV50 - 1	ICV50	15343	3.8376	-7470A	PPB		15 May 2020 14:09:04	U	SMPL
ICB50 - 1	ICB50	378	-0.0199	-7470A	PPB		15 May 2020 14:11:09	U	SMPL
CCV48 - 1	CCV48	18976	4.7741	-7470A	PPB		15 May 2020 14:13:14	U	SMPL
CCB48 - 1	CCB48	387	-0.0176	-7470A	PPB		15 May 2020 14:15:20	U	SMPL
CRA - 1	CRA	1162	0.1822	-7470A	PPB		15 May 2020 14:17:25	U	SMPL
HighStd - 1	HighStd	38890	9.9074	-7470A	PPB		15 May 2020 14:19:31	U	SMPL
ChkStd - 1	ChkStd	27122	6.8739	-7470A	PPB		15 May 2020 14:21:37	U	SMPL
PB128914BL - 1	PBW	195	-0.0671	-7470A	PPB		15 May 2020 14:23:45	U	SMPL
PB128914BS - 1	LCSW	14797	3.6969	-7470A	PPB		15 May 2020 14:25:52	U	SMPL
L2641-05 - 1	OIL-WATER-SEPERATOR-2	238	-0.0560	-7470A	PPB		15 May 2020 14:27:59	U	SMPL
L2644-06 - 1	DRUM-COMP	377	-0.0202	-7470A	PPB		15 May 2020 14:30:06	U	SMPL
L2648-02 - 1	TP-B	311	-0.0372	-7470A	PPB		15 May 2020 14:32:14	U	SMPL
L2648-02DUP - 1	TP-BDUP	377	-0.0202	-7470A	PPB		15 May 2020 14:34:22	U	SMPL
L2648-02MS - 1	TP-BMS	15283	3.8221	-7470A	PPB		15 May 2020 14:36:27	U	SMPL
CCV49 - 1	CCV49	19249	4.8445	-7470A	PPB		15 May 2020 14:38:33	U	SMPL
CCB49 - 1	CCB49	368	-0.0225	-7470A	PPB		15 May 2020 14:40:40	U	SMPL
L2648-02MSD - 1	TP-BMSD	15350	3.8394	-7470A	PPB		15 May 2020 14:42:46	U	SMPL
L2663-05 - 1	1378	239	-0.0558	-7470A	PPB		15 May 2020 14:44:52	U	SMPL
L2663-06 - 1	1381	395	-0.0156	-7470A	PPB		15 May 2020 14:46:58	U	SMPL
L2663-07 - 1	1369	404	-0.0132	-7470A	PPB		15 May 2020 14:49:04	U	SMPL
L2663-08 - 1	1366	350	-0.0272	-7470A	PPB		15 May 2020 14:51:11	U	SMPL
L2663-12 - 1	1359	486	0.0079	-7470A	PPB		15 May 2020 14:53:18	U	SMPL
PB128915BL - 1	PBW	302	-0.0395	-7470A	PPB		15 May 2020 14:55:26	U	SMPL
PB128915BS - 1	LCSW	14725	3.6783	-7470A	PPB		15 May 2020 14:57:33	U	SMPL
L2503-10 - 1	OR-02-051420	283	-0.0444	-7470A	PPB		15 May 2020 14:59:41	U	SMPL
L2503-10DUP - 1	OR-02-051420DUP	284	-0.0442	-7470A	PPB		15 May 2020 15:01:47	U	SMPL
CCV50 - 1	CCV50	19198	4.8313	-7470A	PPB		15 May 2020 15:03:54	U	SMPL
CCB50 - 1	CCB50	394	-0.0158	-7470A	PPB		15 May 2020 15:06:00	U	SMPL
L2503-10MS - 1	OR-02-051420MS	14833	3.7062	-7470A	PPB		15 May 2020 15:08:06	U	SMPL
L2503-10MSD - 1	OR-02-051420MSD	14354	3.5827	-7470A	PPB		15 May 2020 15:10:12	U	SMPL
L2646-04 - 1	TP-B-S	200	-0.0658	-7470A	PPB		15 May 2020 15:12:19	U	SMPL
L2646-08 - 1	TP-B-D	267	-0.0485	-7470A	PPB		15 May 2020 15:14:25	U	SMPL
L2663-04 - 1	72-11935	245	-0.0542	-7470A	PPB		15 May 2020 15:16:32	U	SMPL
L2663-09 - 1	1364	424	-0.0081	-7470A	PPB		15 May 2020 15:18:39	U	SMPL
L2663-10 - 1	1363	315	-0.0362	-7470A	PPB		15 May 2020 15:20:46	U	SMPL
L2663-11 - 1	1328	299	-0.0403	-7470A	PPB		15 May 2020 15:22:53	U	SMPL
PB128916BL - 1	PBW	340	-0.0297	-7470A	PPB		15 May 2020 15:25:01	U	SMPL
PB128916BS - 1	LCSW	14043	3.5025	-7470A	PPB		15 May 2020 15:27:07	U	SMPL
CCV51 - 1	CCV51	18788	4.7256	-7470A	PPB		15 May 2020 15:29:14	U	SMPL
CCB51 - 1	CCB51	58	-0.1024	-7470A	PPB		15 May 2020 15:31:20	U	SMPL
L2648-01 - 1	TP-B	677	0.0571	-7470A	PPB		15 May 2020 15:35:59	U	SMPL
L2648-01DUP - 1	TP-BDUP	546	0.0234	-7470A	PPB		15 May 2020 15:38:05	U	SMPL
L2648-01MS - 1	TP-BMS	15031	3.7572	-7470A	PPB		15 May 2020 15:40:11	U	SMPL
L2648-01MSD - 1	TP-BMSD	16493	4.1341	-7470A	PPB		15 May 2020 15:42:17	U	SMPL
L2665-04 - 1	0373-FIELD-BLANK	117	-0.0872	-7470A	PPB		15 May 2020 15:44:24	U	SMPL
PB128914TB - 1	PB128894TB	325	-0.0336	-7470A	PPB		15 May 2020 15:46:31	U	SMPL
PB128915TB - 1	PB128895TB	344	-0.0287	-7470A	PPB		15 May 2020 15:48:38	U	SMPL
CCV52 - 1	CCV52	18790	4.7262	-7470A	PPB		15 May 2020 15:50:45	U	SMPL
CCB52 - 1	CCB52	97	-0.0924	-7470A	PPB		15 May 2020 15:52:53	U	SMPL
L2648-02LX5 - 1		443	-0.0032	-7470A	PPB		15 May 2020 15:54:59	U	SMPL
L2648-02A - 1		14791	3.6953	-7470A	PPB		15 May 2020 15:57:05	U	SMPL
L2503-10LX5 - 1		377	-0.0202	-7470A	PPB		15 May 2020 15:59:12	U	SMPL
L2503-10A - 1		15014	3.7528	-7470A	PPB		15 May 2020 16:01:18	U	SMPL
L2648-01LX5 - 1		388	-0.0174	-7470A	PPB		15 May 2020 16:03:24	U	SMPL
L2648-01A - 1		15350	3.8394	-7470A	PPB		15 May 2020 16:05:31	U	SMPL
CCV53 - 1	CCV53	18931	4.7625	-7470A	PPB		15 May 2020 16:07:37	U	SMPL
CCB53 - 1	CCB53	380	-0.0194	-7470A	PPB		15 May 2020 16:09:43	U	SMPL