



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109198

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/15/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.01	05/15/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	05/15/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.00	05/15/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	05/15/2020	17:16
6	L2673-01	1	Water	NA	NA	24.4	5.89	05/15/2020	17:35
7	L2673-01DUP	1	Water	NA	NA	24.5	5.90	05/15/2020	17:36
8	L2673-03	1	Water	NA	NA	24.5	7.27	05/15/2020	17:39
9	CCV2	1	Water	NA	NA	20.2	12.02	05/15/2020	17:40

WORKLIST(Hardcopy Internal Chain)

183109198

WorkList Name : PH L2573

WorkList ID : 138798

Department : Wet-Chemistry

Date : 05-15-2020 16:00:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/15/2020	Water	L2673-01	pH	Cool 4 deg C	PSEG01	M51	102	05/15/2020	9040C
05/15/2020	Water	L2673-03	pH	Cool 4 deg C	PSEG01	M51	104	05/15/2020	9040C

Date/Time 05.15.2020 16:10

Received by: JP

Relinquished by: KS

Date/Time 05.13.2020 18:15

Received by: KS

Relinquished by: JP