



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109206

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/18/2020	10:25
2	CAL2	1	Water	NA	NA	20.1	7.00	05/18/2020	10:26
3	CAL3	1	Water	NA	NA	20.1	10.03	05/18/2020	10:27
4	ICV	1	Water	NA	NA	20.2	7.00	05/18/2020	10:30
5	CCV1	1	Water	NA	NA	20.1	2.01	05/18/2020	10:31
6	L2685-01	1	Water	NA	NA	20.1	5.19	05/18/2020	10:35
7	L2685-01DUP	1	Water	NA	NA	20.2	5.20	05/18/2020	10:36
8	CCV2	1	Water	NA	NA	20.1	12.02	05/18/2020	10:37

VB 109206

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2685 WorkList ID : 138835 Department : Wet-Chemistry Date : 05-18-2020 10:15:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/13/2020	Water	L2685-01-17	pH	Cool 4 deg C	STER08	N11	POOL-WATER	05/13/2020	SM 4500-H B

Date/Time 05/18/2020 10:20
Received by: Jo
Relinquished by: Jo

Date/Time 05/18/2020 11:20
Received by: Jo
Relinquished by: Jo