


LB109207

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1102e-004

C= -7.5134e-002

Rho= 0.9999916

Accept=Accepted

slope
y=10715.037x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	326	0.000	326					
0.05	0.050					223					
0.20	0.200	0.191	-0.009	1261	0.0 %	1261					
2.50	2.500	2.509	0.009	12247	0.0 %	12247					
5.00	5.000	5.027	0.027	24180	0.0 %	24180					
7.50	7.500	7.479	-0.021	35800	0.0 %	35800					
10.0	10.000	10.000	-0.000	47744	0.0 %	47744					

9/10
= 6.5
0 5/18/20

LB109207
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	326	-		0 ISM02.4	PPB	5/18/2020 12:35	Std
	0.2 S0.2	1261	-		0.2 ISM02.4	PPB	5/18/2020 12:37	Std
	2.5 S2.5	12247	-		2.5 ISM02.4	PPB	5/18/2020 12:40	Std
	5 S5.0	24180	-		5 ISM02.4	PPB	5/18/2020 12:42	Std
	7.5 S7.5	35800	-		7.5 ISM02.4	PPB	5/18/2020 12:44	Std
	10 S10.0	47744	-		10 ISM02.4	PPB	5/18/2020 12:46	Std
ICV	ICV	18791	3.8901 -		ISM02.4	PPB	5/18/2020 12:49	SMPL
ICB	ICB	203	-0.0323 -		ISM02.4	PPB	5/18/2020 12:51	SMPL
CCV057	CCV057	24306	5.0538 -		ISM02.4	PPB	5/18/2020 12:54	SMPL
CCB057	CCB057	213	-0.0302 -		ISM02.4	PPB	5/18/2020 12:56	SMPL
PB128939BL	PBS939	58	-0.0629 -		ISM02.4	PPB	5/18/2020 12:58	SMPL
L2651-01	MH0KN9	673	0.0669 -		ISM02.4	PPB	5/18/2020 13:00	SMPL
L2651-02	MH0KP0	453	0.0205 -		ISM02.4	PPB	5/18/2020 13:03	SMPL
L2651-03	MH0KP1	636	0.0591 -		ISM02.4	PPB	5/18/2020 13:05	SMPL
L2651-04	MH0KP2	4502	0.8749 -		ISM02.4	PPB	5/18/2020 13:07	SMPL
L2651-05	MH0KP3	2562	0.4655 -		ISM02.4	PPB	5/18/2020 13:09	SMPL
L2651-06	MH0KP4	10894	2.2237 -		ISM02.4	PPB	5/18/2020 13:12	SMPL
L2651-07	MH0KP5	3533	0.6704 -		ISM02.4	PPB	5/18/2020 13:14	SMPL
L2651-08	MH0KP6	2773	0.51 -		ISM02.4	PPB	5/18/2020 13:16	SMPL
L2651-09	MH0KP7	555	0.042 -		ISM02.4	PPB	5/18/2020 13:18	SMPL
L2651-10	MH0KP8	470	0.024 -		ISM02.4	PPB	5/18/2020 13:21	SMPL
L2651-11	MH0KP9	482	0.0266 -		ISM02.4	PPB	5/18/2020 13:23	SMPL
L2651-12	MH0KQ0	1341	0.2078 -		ISM02.4	PPB	5/18/2020 13:25	SMPL
L2651-13	MH0KQ1	923	0.1196 -		ISM02.4	PPB	5/18/2020 13:27	SMPL
L2651-14	MH0KQ2	201	-0.0327 -		ISM02.4	PPB	5/18/2020 13:30	SMPL
L2651-15	MH0KQ3	262	-0.0198 -		ISM02.4	PPB	5/18/2020 13:32	SMPL
L2651-16	MH0KQ4	90	-0.0561 -		ISM02.4	PPB	5/18/2020 13:34	SMPL
L2651-17	MH0KQ5	177	-0.0378 -		ISM02.4	PPB	5/18/2020 13:37	SMPL
L2651-18	MH0KQ6	384	0.0059 -		ISM02.4	PPB	5/18/2020 13:39	SMPL
L2651-19	MH0KQ7	184	-0.0363 -		ISM02.4	PPB	5/18/2020 13:41	SMPL
CCV058	CCV058	24119	5.0144 -		ISM02.4	PPB	5/18/2020 13:43	SMPL
CCB058	CCB058	192	-0.0346 -		ISM02.4	PPB	5/18/2020 13:46	SMPL
L2651-20	MH0KQ8	426	0.0148 -		ISM02.4	PPB	5/18/2020 13:48	SMPL
L2651-21	MH0KQ8D	374	0.0038 -		ISM02.4	PPB	5/18/2020 13:50	SMPL
L2651-22	MH0KQ8S	11584	2.3693 -		ISM02.4	PPB	5/18/2020 13:52	SMPL
PB128940BL	PBS940	92	-0.0557 -		ISM02.4	PPB	5/18/2020 13:55	SMPL
L2652-01	MH0KQ9	240	-0.0245 -		ISM02.4	PPB	5/18/2020 13:57	SMPL
L2652-02	MH0KR0	1455	0.2319 -		ISM02.4	PPB	5/18/2020 13:59	SMPL
L2652-03	MH0KR1	1944	0.3351 -		ISM02.4	PPB	5/18/2020 14:06	SMPL
L2652-04	MH0KR1D	1969	0.3404 -		ISM02.4	PPB	5/18/2020 14:09	SMPL
L2652-05	MH0KR1S	14256	2.9331 -		ISM02.4	PPB	5/18/2020 14:11	SMPL
PB128941BL	PBS941	-24	-0.0802 -		ISM02.4	PPB	5/18/2020 14:13	SMPL
L2659-01	MCB7C0	411	0.0116 -		ISM02.4	PPB	5/18/2020 14:16	SMPL
L2659-02	MCB7C1	910	0.1169 -		ISM02.4	PPB	5/18/2020 14:18	SMPL
L2659-03	MCB7C2	1217	0.1817 -		ISM02.4	PPB	5/18/2020 14:20	SMPL
L2659-04	MCB7C3	819	0.0977 -		ISM02.4	PPB	5/18/2020 14:22	SMPL

LB109207

INSTRUMENT ID : CV1

L2659-05	MCB7C3D	835	0.1011 -	ISM02.4	PPB	5/18/2020 14:27	SMPL
L2659-06	MCB7C3S	12823	2.6307 -	ISM02.4	PPB	5/18/2020 14:29	SMPL
CCV059	CCV059	24348	5.0627 -	ISM02.4	PPB	5/18/2020 14:32	SMPL
CCB059	CCB059	190	-0.035 -	ISM02.4	PPB	5/18/2020 14:34	SMPL
L2659-07	MCB7C4	320	-0.0076 -	ISM02.4	PPB	5/18/2020 14:36	SMPL
L2659-08	MCB7C5	290	-0.0139 -	ISM02.4	PPB	5/18/2020 14:38	SMPL
L2659-09	MCB7C6	1048	0.146 -	ISM02.4	PPB	5/18/2020 14:41	SMPL
L2659-10	MCB7C7	441	0.0179 -	ISM02.4	PPB	5/18/2020 14:43	SMPL
L2659-11	MCB7C8	3905	0.7489 -	ISM02.4	PPB	5/18/2020 14:45	SMPL
L2659-12	MCB7C9	5882	1.1661 -	ISM02.4	PPB	5/18/2020 14:48	SMPL
L2659-13	MCB7D0	1442	0.2292 -	ISM02.4	PPB	5/18/2020 14:50	SMPL
L2659-14	MCB7D1	4111	0.7924 -	ISM02.4	PPB	5/18/2020 14:52	SMPL
CCV060	CCV060	24388	5.0711 -	ISM02.4	PPB	5/18/2020 14:54	SMPL
CCB060	CCB060	196	-0.0338 -	ISM02.4	PPB	5/18/2020 14:57	SMPL