



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109211

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	05/18/2020	13:30
2	CAL2	1	Water	NA	NA	20.1	7.01	05/18/2020	13:31
3	CAL3	1	Water	NA	NA	20.1	10.03	05/18/2020	13:32
4	ICV	1	Water	NA	NA	20.1	7.00	05/18/2020	13:35
5	CCV1	1	Water	NA	NA	20.2	2.01	05/18/2020	13:36
6	L2687-01	1	Solid	20.03	20	22.2	11.25	05/18/2020	13:37
7	L2687-01DUP	1	Solid	20.04	20	22.3	11.26	05/18/2020	13:38
8	CCV2	1	Water	NA	NA	20.1	12.02	05/18/2020	13:40

VB 169211

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2687 WorkList ID : 138841 Department : Wet-Chemistry Date : 05-18-2020 13:22:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/18/2020	Solid	L2687-01	pH	Cool 4 deg C	SEAC01	M51	MUD	05/18/2020	9045D

Date/Time 05/18/2020 13:25
Received by: JP
Relinquished by: JP

Date/Time 05/18/2020 15:35
Received by: JP
Relinquished by: JP