



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109216

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/18/2020	09:20
2	L2673-01		1	8	100.00	>212.0	>212.0	05/18/2020	09:50
3	L2673-01DUP		1	8	100.00	>212.0	>212.0	05/18/2020	10:15
4	L2673-02		1	9	100.00	>212.0	>212.0	05/18/2020	10:45
5	L2673-03		1	7	100.00	>212.0	>212.0	05/18/2020	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp51820 WorkList ID : 138828 Department : Wet-Chemistry Date : 05-18-2020 08:40:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2020	Water	L2673-01	Flash Point	Cool 4 deg C	PSEG01	M51	102	05/15/2020	1010A
05/22/2020	Water	L2673-02	Flash Point	Cool 4 deg C	PSEG01	M51	30559	05/15/2020	1010A
05/22/2020	Water	L2673-03	Flash Point	Cool 4 deg C	PSEG01	M51	104	05/15/2020	1010A

Date/Time 05/18/20 08:45
 Received by: KS
 Relinquished by: JP

Date/Time 05/18/20 11:25
 Received by: JP
 Relinquished by: KS