


LB109225

ISM02.4

INSTRUMENT ID: CY1



Linear

A= 0.0000e+000

B= 2.2609e-004

C= 2.0485e-002

Rho= 0.9999101

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.027	0.027	30	0.000	30					11
0.05	0.050					223					3
0.20	0.200	0.206	0.006	820	0.0 %	820					1
2.50	2.500	2.530	0.030	11099	0.0 %	11099					2
5.00	5.000	4.902	-0.098	21589	0.0 %	21589					0
7.50	7.500	7.481	-0.019	32996	0.0 %	32996					0
10.0	10.000	10.055	0.055	44380	0.0 %	44380					1

LB109225

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	30	-		0 ISM02.4	PPB	5/19/2020 13:59	Std
	0.2 S0.2	820	-		0.2 ISM02.4	PPB	5/19/2020 14:01	Std
	2.5 S2.5	11099	-		2.5 ISM02.4	PPB	5/19/2020 14:03	Std
	5 S5.0	21589	-		5 ISM02.4	PPB	5/19/2020 14:05	Std
	7.5 S7.5	32996	-		7.5 ISM02.4	PPB	5/19/2020 14:08	Std
	10 S10.0	44380	-		10 ISM02.4	PPB	5/19/2020 14:10	Std
ICV	ICV	16514	3.7542 -		ISM02.4	PPB	5/19/2020 14:13	SMPL
ICB	ICB	-73	0.004 -		ISM02.4	PPB	5/19/2020 14:15	SMPL
CCV066	CCV066	21617	4.908 -		ISM02.4	PPB	5/19/2020 14:17	SMPL
CCB066	CCB066	-66	0.0056 -		ISM02.4	PPB	5/19/2020 14:19	SMPL
PB128993BL	PBW993	1	0.0207 -		ISM02.4	PPB	5/19/2020 14:22	SMPL
L2654-01	MC0AE0	-128	-0.0085 -		ISM02.4	PPB	5/19/2020 14:24	SMPL
L2654-02	MC0AE1	-138	-0.0107 -		ISM02.4	PPB	5/19/2020 14:26	SMPL
L2654-03	MC0AD8	-121	-0.0069 -		ISM02.4	PPB	5/19/2020 14:28	SMPL
L2654-04	MC0AD9	-119	-0.0064 -		ISM02.4	PPB	5/19/2020 14:31	SMPL
L2654-05	MC0AD9D	-121	-0.0069 -		ISM02.4	PPB	5/19/2020 14:33	SMPL
L2654-06	MC0AD9S	4295	0.9916 -		ISM02.4	PPB	5/19/2020 14:35	SMPL
CCV067	CCV067	21179	4.8089 -		ISM02.4	PPB	5/19/2020 14:51	SMPL
CCB067	CCB067	-134	-0.0098 -		ISM02.4	PPB	5/19/2020 14:53	SMPL