


LB109226

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1490e-004

C= -3.0167e-002

Rho= 0.9999369

Accept=Accepted

slope
 $y = 1.5716 \times 10^{-4} x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.021	-0.021	43	0.000	43					
0.05	0.050					223					
0.20	0.200	0.173	-0.027	944	0.0 %	944					
2.50	2.500	2.577	0.077	12132	0.0 %	12132					
5.00	5.000	5.009	0.009	23447	0.0 %	23447					
7.50	7.500	7.446	-0.054	34791	0.0 %	34791					
10.0	10.000	10.016	0.016	46748	0.0 %	46748					

70D
 14
 3
 0
 10

LB109226

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	43	-	0	ISM02.4	PPB	5/19/2020 15:15	Std
	0.2 S0.2	944	-	0.2	ISM02.4	PPB	5/19/2020 15:17	Std
	2.5 S2.5	12132	-	2.5	ISM02.4	PPB	5/19/2020 15:20	Std
	5 S5.0	23447	-	5	ISM02.4	PPB	5/19/2020 15:22	Std
	7.5 S7.5	34791	-	7.5	ISM02.4	PPB	5/19/2020 15:24	Std
	10 S10.0	46748	-	10	ISM02.4	PPB	5/19/2020 15:28	Std
ICV	ICV	18337	3.9105 -		ISM02.4	PPB	5/19/2020 15:31	SMPL
ICB	ICB	-83	-0.048 -		ISM02.4	PPB	5/19/2020 15:33	SMPL
CCV068	CCV068	23607	5.043 -		ISM02.4	PPB	5/19/2020 15:35	SMPL
CCB068	CCB068	-105	-0.0527 -		ISM02.4	PPB	5/19/2020 15:38	SMPL
PB128995BL	PBS995	-5	-0.0312 -		ISM02.4	PPB	5/19/2020 15:40	SMPL
L2667-19	MX2M43	9182	1.9431 -		ISM02.4	PPB	5/19/2020 15:44	SMPL
L2667-20	MX2M43D	9564	2.0252 -		ISM02.4	PPB	5/19/2020 15:47	SMPL
L2667-21	MX2M43S	20591	4.3949 -		ISM02.4	PPB	5/19/2020 15:49	SMPL
CCV069	CCV069	22771	4.8634 -		ISM02.4	PPB	5/19/2020 15:51	SMPL
CCB069	CCB069	-94	-0.0504 -		ISM02.4	PPB	5/19/2020 15:53	SMPL