



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109227

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/19/2020	14:30
2	CAL2	1	Water	NA	NA	20.1	7.00	05/19/2020	14:31
3	CAL3	1	Water	NA	NA	20.2	10.03	05/19/2020	14:32
4	ICV	1	Water	NA	NA	20.2	7.02	05/19/2020	14:35
5	CCV1	1	Water	NA	NA	20.1	2.01	05/19/2020	14:36
6	L2714-01	1	Water	NA	NA	16.9	7.00	05/19/2020	14:37
7	L2714-01DUP	1	Water	NA	NA	16.8	7.01	05/19/2020	14:38
8	CCV2	1	Water	NA	NA	20.1	12.02	05/19/2020	14:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2714 WorkList ID : 138885 Department : Wet-Chemistry Date : 05-19-2020 14:24:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/19/2020	Water	L2714-01-C	pH	Cool 4 deg C	PSEG01	N12	RP-WATER	05/19/2020	9040C

Date/Time 05.19.2020 14:15
 Received by: JP
 Relinquished by: JP

Date/Time 05.19.2020 15:30
 Received by: JP
 Relinquished by: JP