



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/20/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:25
In Date: 05/19/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 05/20/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB109230

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2497-05	HD-02-051920	1	1.14	9.62	9.24	95.5	
L2497-06	HD-02-051920	2	1.13	9.79	9.02	91.1	
L2497-07	HD-02-051920-1	3	1.15	9.71	9.6	98.7	
L2692-01	400512005.0	4	1.15	9.9	1.74	6.7	sludge sample
L2693-01	400513009.0	5	1.19	9.97	1.68	5.6	sludge sample
L2694-01	400513712.0	6	1.14	9.94	2.73	18.1	sludge sample
L2695-01	400513704.2	7	1.13	9.96	1.54	4.6	sludge sample
L2696-01	400513720.0	8	1.13	9.78	3.44	26.7	SLUDGE SAMPLE
L2697-01	400514719.2	9	1.19	9.8	1.45	3.0	sludge sample
L2698-01	400513118.1	10	1.17	9.98	1.62	5.1	sludge sample
L2710-01	PCB-051420-SR-01	11	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-02	PCB-051420-SR-02	12	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-03	PCB-051420-SR-03	13	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-04	PCB-051420-SR-04	14	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-05	PCB-051420-SR-05	15	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-06	PCB-051420-SR-06	16	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-07	PCB-051420-SR-07	17	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-08	PCB-051420-SR-08	18	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2710-09	PCB-051420-SR-09	19	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$