



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109231

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/19/2020	16:30
2	CAL2	1	Water	NA	NA	20.1	7.00	05/19/2020	16:31
3	CAL3	1	Water	NA	NA	20.1	10.04	05/19/2020	16:32
4	ICV	1	Water	NA	NA	20.2	6.98	05/19/2020	16:33
5	CCV1	1	Water	NA	NA	20.2	2.01	05/19/2020	16:35
6	L2497-08	1	Solid	20.03	20	22.8	8.57	05/19/2020	16:36
7	L2497-08DUP	1	Solid	20.04	20	22.9	8.58	05/19/2020	16:37
8	L2713-01	1	Solid	20.01	20	20.5	6.10	05/19/2020	16:30
9	L2713-02	1	Solid	20.03	20	21.1	6.18	05/19/2020	16:32
10	CCV2	1	Water	NA	NA	20.1	12.02	05/19/2020	16:33

WORKLIST(Hardcopy Internal Chain)

LB 109231

WorkList Name : CORROSIVITY L2713 WorkList ID : 138884 Department : Wet-Chemistry Date : 05-19-2020 13:17:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/19/2020	Solid	L2497-08	Corrosivity	Cool 4 deg C	PSEG05		HD-02-051920	05/19/2020	9045D
05/18/2020	Solid	L2713-01	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH-SLUDGE-PIT	05/18/2020	9045D
05/18/2020	Solid	L2713-02	Corrosivity	Cool 4 deg C	CLEA11		ZIPPER-DRAINS-SLUDGE-PI	05/18/2020	9045D

Date/Time 05-19-2020 15:50
 Received by: JP
 Relinquished by: JP

Date/Time 05-19-2020 17:00
 Received by: JP
 Relinquished by: JP