



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109232

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/19/2020	16:20
2	L2714-01		1	8	100.00	>212.0	>212.0	05/19/2020	16:45
3	L2714-01DUP		1	8	100.00	>212.0	>212.0	05/19/2020	17:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp51920 WorkList ID : 138891 Department : Wet-Chemistry Date : 05-19-2020 15:52:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/26/2020	Water	L2714-01 H	Flash Point	Cool 4 deg C	PSEG01	N12	RP-WATER	05/19/2020	1010A

Date/Time 05/19/20 15:55
 Received by: KS
 Relinquished by: JP

Date/Time 05/19/20 19:15
 Received by: JP
 Relinquished by: KS