

LB109241


7471B

INSTRUMENT 1D: CV1

Linear

 μ Abs.:

52840



A= 0.0000e+000

B= 1.8789e-004

C= -7.9035e-002

Rho= 0.9995757

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	206	0.000	206				
0.2	0.200	0.238	0.038	1686	0.0 %	1686				
2.5	2.500	2.410	-0.090	13245	0.0 %	13245				
5.0	5.000	5.069	0.069	27396	0.0 %	27396				
7.5	7.500	7.675	0.175	41269	0.0 %	41269				
10.0	10.000	9.849	-0.151	52840	0.0 %	52840				

LB109241 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 May 2020 12:00:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	206	-	0.0000	7471B	PPB	20 May 2020 12:02:28	S	Std
0.2 - 1	50.2	1686	-	0.2000	7471B	PPB	20 May 2020 12:04:45	S	Std
2.5 - 1	52.5	13245	-	2.5000	7471B	PPB	20 May 2020 12:07:02	S	Std
5.0 - 1	55	27396	-	5.0000	7471B	PPB	20 May 2020 12:09:18	S	Std
7.5 - 1	57.5	41269	-	7.5000	7471B	PPB	20 May 2020 12:11:34	S	Std
10.0 - 1	510	52840	-	10.0000	7471B	PPB	20 May 2020 12:13:51	S	Std
ICV54 - 1	ICV54	21674	3.9934	-	7471B	PPB	20 May 2020 12:16:45	U	SMPL
ICB54 - 1	ICB54	176	-0.0460	-	7471B	PPB	20 May 2020 12:19:00	U	SMPL
CCV73 - 1	CCV73	27513	5.0905	-	7471B	PPB	20 May 2020 12:21:16	U	SMPL
CCB73 - 1	CCB73	166	-0.0478	-	7471B	PPB	20 May 2020 12:23:30	U	SMPL
CRA - 1	CRA	1654	0.2317	-	7471B	PPB	20 May 2020 12:25:46	U	SMPL
HighStd - 1	HighStd	52786	9.8391	-	7471B	PPB	20 May 2020 12:28:01	U	SMPL
ChkStd - 1	ChkStd	39713	7.3828	-	7471B	PPB	20 May 2020 12:30:17	U	SMPL
PB129007BL - 1	PBS	-51	-0.0886	-	7471B	PPB	20 May 2020 12:35:07	U	SMPL
PB129007BS - 1	LCSS	19602	3.6041	-	7471B	PPB	20 May 2020 12:37:22	U	SMPL
L2497-05 - 1	HD-02-051920	3462	0.5715	-	7471B	PPB	20 May 2020 12:39:38	U	SMPL
L2686-02 - 1	VNJ-214	6399	1.1233	-	7471B	PPB	20 May 2020 12:41:55	U	SMPL
L2686-05 - 1	VNJ-241	10439	1.8824	-	7471B	PPB	20 May 2020 12:44:12	U	SMPL
L2686-05DUP - 1	VNJ-241DUP	10533	1.9001	-	7471B	PPB	20 May 2020 12:46:29	U	SMPL
L2686-05MS - 1	VNJ-241MS	31423	5.8252	-	7471B	PPB	20 May 2020 12:48:44	U	SMPL
CCV74 - 1	CCV74	27185	5.0289	-	7471B	PPB	20 May 2020 12:50:59	U	SMPL
CCB74 - 1	CCB74	299	-0.0229	-	7471B	PPB	20 May 2020 12:55:45	U	SMPL
L2686-05MSD - 1	VNJ-241MSD	30294	5.6130	-	7471B	PPB	20 May 2020 12:57:59	U	SMPL
L2686-05LX5 - 1		2161	0.3270	-	7471B	PPB	20 May 2020 13:00:14	U	SMPL
L2686-05A - 1		30865	5.7203	-	7471B	PPB	20 May 2020 13:02:30	U	SMPL
CCV75 - 1	CCV75	27437	5.0762	-	7471B	PPB	20 May 2020 13:04:46	U	SMPL
CCB75 - 1	CCB75	185	-0.0443	-	7471B	PPB	20 May 2020 13:07:03	U	SMPL