


LB109246

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2882e-004

C= -1.4195e-002

Rho= 0.9999121

Accept=Accepted

slope
y=10000x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.001	0.001	65	0.000	65					
0.05	0.050					223					
0.20	0.200	0.195	-0.005	916	0.0 %	916					
2.50	2.500	2.525	0.025	11095	0.0 %	11095					
5.00	5.000	5.024	0.024	22018	0.0 %	22018					
7.50	7.500	7.400	-0.100	32401	0.0 %	32401					
10.0	10.000	10.056	0.056	44009	0.0 %	44009					

100
-3
-1
-0
-1
-1

LB109246

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	65	-		0 ISM02.4	PPB	5/20/2020 14:01	Std
	0.2 S0.2	916	-		0.2 ISM02.4	PPB	5/20/2020 14:03	Std
	2.5 S2.5	11095	-		2.5 ISM02.4	PPB	5/20/2020 14:05	Std
	5 S5.0	22018	-		5 ISM02.4	PPB	5/20/2020 14:08	Std
	7.5 S7.5	32401	-		7.5 ISM02.4	PPB	5/20/2020 14:10	Std
	10 S10.0	44009	-		10 ISM02.4	PPB	5/20/2020 14:12	Std
ICV	ICV	16814	3.8331 -		ISM02.4	PPB	5/20/2020 14:15	SMPL
ICB	ICB	-78	-0.032 -		ISM02.4	PPB	5/20/2020 14:17	SMPL
CCV076	CCV076	22025	5.0255 -		ISM02.4	PPB	5/20/2020 14:20	SMPL
CCB076	CCB076	-81	-0.0327 -		ISM02.4	PPB	5/20/2020 14:22	SMPL
PB129025BL	PBW025	-19	-0.0185 -		ISM02.4	PPB	5/20/2020 14:24	SMPL
L2667-07	MX2M39	39039	8.9186 -		ISM02.4	PPB	5/20/2020 14:26	SMPL
L2667-08	MX2M39D	39193	8.9538 -		ISM02.4	PPB	5/20/2020 14:31	SMPL
L2667-09	MX2M39S	43705	9.9862 -		ISM02.4	PPB	5/20/2020 14:44	SMPL
CCV077	CCV077	21684	4.9475 -		ISM02.4	PPB	5/20/2020 14:46	SMPL
CCB077	CCB077	-88	-0.0343 -		ISM02.4	PPB	5/20/2020 14:48	SMPL